

11/4/08

OFF01-110408 (N of Warm Springs)

TSP #002 run 757 min ^{TS} 13.13 M run complete bath 52%

Abbst. #0097 run 791 min 2L/min

T04 00314.67 56

T09 00298.30 52

T013 00371.51 48 end 750 AM

OFF02-110408 (S of Warm Springs)

TSP - #11073 830 min ^T 13.97 M bath 58%

Abbst. - #110154 830 min 2L/min

T04 - 48 00318.65

T09 - 52 00295.96

T013 - 44 00405.50 end 811 AM

TETRA TECH EM INC.
AIR QUALITY SAMPLE DOCUMENTATION SHEET
BMI COMPLEX SITE
SITE INFORMATION

Daily Temperature: 59°F

Beginning Ambient Baro. Press. (in. Hg): 30.28 Ending Ambient Baro. Pressure (in Hg.) 30.24

Sample Location OFF01- 110708

Start Time: 1830 End Time: 0659

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1075

Beginning Pressure Drop (Inches H2O): 62

Beginning Timer Reading (hours): 00314.67

Ending Pressure Drop (Inches H2O): 59

Ending Timer Reading (hours): 00324.55 11.88 hr

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1077

Beginning Pressure Drop (Inches H2O): 80

Beginning Timer Reading (hours): 00298.37

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00310.27 11.90 hr

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1074

Beginning Pressure Drop (Inches H2O): 100

Beginning Timer Reading (hours): 00371.52

Ending Pressure Drop (Inches H2O): 59

Ending Timer Reading (hours): 00383.41 11.89 hr

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 08T6433

Beginning Flow Rate (LPM): 16.7 Battery % 99/79

Timer Beginning Time: 0

Ending Flow Rate (LPM): 16.7

Timer Ending Time: 760

BGI PQ100 Sampler Serial #: 11673

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 760

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 00997

Sample Location OFF02- 110708

Start Time: 1852 End Time: 0715

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1079

Beginning Pressure Drop (Inches H2O): 59

Beginning Timer Reading (hours): 00318.45

Ending Pressure Drop (Inches H2O): 54

Ending Timer Reading (hours): 00327.48 8.83 hr

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1076

Beginning Pressure Drop (Inches H2O): 59

Beginning Timer Reading (hours): 00295.96

Ending Pressure Drop (Inches H2O): 52

Ending Timer Reading (hours): 00308.09 12.13 hr

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1078

Beginning Pressure Drop (Inches H2O): 59

Beginning Timer Reading (hours): 00405.50

Ending Pressure Drop (Inches H2O): 52

Ending Timer Reading (hours): 00417.42 12.12 hr

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 08T6434

Beginning Flow Rate (LPM): 16.7 Battery % 58/12

Timer Beginning Time: 0

Ending Flow Rate (LPM): 16.7

Timer Ending Time: 410

BGI PQ100 Sampler Serial #: 002 low batt. shut down

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 760

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 10154

Comments: Pump 002 did not take full charge. Called Doug to let him know to call Pine for replacement pump.

Refill generators 2AM-1AM ? unsure why T04 on OFF2 time is short ?
Signature: [Signature] timer #3 written down correctly. did not shut off

TETRA TECH EM INC.
AIR QUALITY SAMPLE DOCUMENTATION SHEET
BMI COMPLEX SITE
SITE INFORMATION

Daily Temperature: 58°F

Beginning Ambient Baro. Press. (in. Hg): 29.97 Ending Ambient Baro. Pressure (In Hg.): 29.94

Sample Location OFF01- 111108

Start Time: 1850 End Time:

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1075

Beginning Pressure Drop (Inches H2O): 59

Beginning Timer Reading (hours): 00326.55

Ending Pressure Drop (Inches H2O): 54

Ending Timer Reading (hours): 00338.02 12.07

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1077

Beginning Pressure Drop (Inches H2O): 56

Beginning Timer Reading (hours): 00310.20

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00322.33 12.13

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1074

Beginning Pressure Drop (Inches H2O): 52

Beginning Timer Reading (hours): 00383.41

Ending Pressure Drop (Inches H2O): 49

Ending Timer Reading (hours): 00395.47 12.00

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 0876442

Beginning Flow Rate (LPM): 16.7 Battery % 99 / 90

Timer Beginning Time: 0

Ending Flow Rate (LPM): 14.7

Timer Ending Time: 0763

BGI PQ100 Sampler Serial #: 11673

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 765

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 10154

Sample Location OFF02- 111108

Start Time: 1910 End Time:

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1079

Beginning Pressure Drop (Inches H2O): 56

Beginning Timer Reading (hours): 00327.48

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00339.73 12.25

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1076

Beginning Pressure Drop (Inches H2O): 60

Beginning Timer Reading (hours): 00308.09

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00320.34 12.25

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1078

Beginning Pressure Drop (Inches H2O): 54

Beginning Timer Reading (hours): 00417.02

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00429.87 12.25

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 0876439

Beginning Flow Rate (LPM): 16.7 Battery % 40 / 10

Timer Beginning Time: 0

Ending Flow Rate (LPM): 16.7

Timer Ending Time: 90

BGI PQ100 Sampler Serial #: 602 Low Batt Shutdown!

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 765

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 2997

Comments: Site OFF02 started late - had to return to trailer to get a different T04 filter. Original filter too wide for sample unit. Field generators between 12AM-1AM

Signature: [Signature]

TETRA TECH EM INC.
AIR QUALITY SAMPLE DOCUMENTATION SHEET
BMI COMPLEX SITE
SITE INFORMATION

Daily Temperature: Low 53°F

Beginning Ambient Baro. Press. (in. Hg): 29.84 Ending Ambient Baro. Pressure (in Hg.): 30.210

Sample Location OFF01- 11/408

Start Time: 1840 End Time: 0720

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1075

Beginning Pressure Drop (Inches H2O): 50

Beginning Timer Reading (hours): 00338.62

Ending Pressure Drop (Inches H2O): 52

Ending Timer Reading (hours): 00350.63

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1077

Beginning Pressure Drop (Inches H2O): 54

Beginning Timer Reading (hours): 00322.33

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00334.34

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1074

Beginning Pressure Drop (Inches H2O): 48

Beginning Timer Reading (hours): 00395.47

Ending Pressure Drop (Inches H2O): 48

Ending Timer Reading (hours): 00407.49

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 0876447

Beginning Flow Rate (LPM): 16.7 Battery % 99/82

Timer Beginning Time: 0

Ending Flow Rate (LPM):

Timer Ending Time: 0758

BGI PQ100 Sampler Serial #: 08710

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 0760

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 10154

Sample Location OFF02- 11/408

Start Time: 1845 End Time: 0735

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1079

Beginning Pressure Drop (Inches H2O): —

Beginning Timer Reading (hours): —

Ending Pressure Drop (Inches H2O): —

Ending Timer Reading (hours): —

**NO SAMPLE
SEE COMMENTS**

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1076

Beginning Pressure Drop (Inches H2O): 58

Beginning Timer Reading (hours): 00320.34

Ending Pressure Drop (Inches H2O): 52

Ending Timer Reading (hours): 00332.85

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1078

Beginning Pressure Drop (Inches H2O): 52

Beginning Timer Reading (hours): 00429.87

Ending Pressure Drop (Inches H2O): 52

Ending Timer Reading (hours): 00442.39

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 0876448

Beginning Flow Rate (LPM): 16.7 Battery % 99/88

Timer Beginning Time: 0

Ending Flow Rate (LPM): 16.7

Timer Ending Time: 0760

BGI PQ100 Sampler Serial #: 11673

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 0760

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 20917

Comments: OFF02 NO TO4 sample. Additional Filter from Air Toxics Dip NOT
ARRIVE TO REPLACE NO FIT FILTER FROM NOV 10-11 sample set.

Signature: [Signature]

TETRA TECH EM INC.
AIR QUALITY SAMPLE DOCUMENTATION SHEET
BMI COMPLEX SITE
SITE INFORMATION

Daily Temperature:

Beginning Ambient Baro. Press. (In. Hg): Ending Ambient Baro. Pressure (in Hg.)

Sample Location OFF01- 111808

Start Time: 1829

End Time: 0645

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1075

Beginning Pressure Drop (Inches H2O): 50

Beginning Timer Reading (hours): 00350.63

Ending Pressure Drop (Inches H2O): 54

Ending Timer Reading (hours): 003102.39

11.70

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1077

Beginning Pressure Drop (Inches H2O): 52

Beginning Timer Reading (hours): 00334.34

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00346.09

11.75

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1074

Beginning Pressure Drop (Inches H2O): 34

Beginning Timer Reading (hours): 00407.49

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00419.84

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 08T6454

Beginning Flow Rate (LPM): 16.7

Battery % 99/86

Timer Beginning Time: 0

Ending Flow Rate (LPM): 16.7

Timer Ending Time: 0745

BGI PQ100 Sampler Serial #: 08710

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 0745

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 2997

Sample Location OFF02- 111808

Start Time: 1850

End Time: 0705

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1079

Beginning Pressure Drop (Inches H2O): 52

Beginning Timer Reading (hours): 00339.73

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00351.74

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1076

Beginning Pressure Drop (Inches H2O): 60

Beginning Timer Reading (hours): 00332.85

Ending Pressure Drop (Inches H2O): 52

Ending Timer Reading (hours): 00344.86

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1078

Beginning Pressure Drop (Inches H2O): 54

Beginning Timer Reading (hours): 00442.39

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00454.39

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 08T6453

Beginning Flow Rate (LPM): 16.7

Battery % 99/91

Timer Beginning Time: 0

Ending Flow Rate (LPM): 16.7

Timer Ending Time: 0745

BGI PQ100 Sampler Serial #: 11673

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 0745

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 10154

Comments: all sample runs complete

Signature: Bethi Lane

TETRA TECH EM INC.
AIR QUALITY SAMPLE DOCUMENTATION SHEET
BMI COMPLEX SITE
SITE INFORMATION

Daily Temperature: 1-20 91°F

Beginning Ambient Baro. Press. (In. Hg): Ending Ambient Baro. Pressure (in Hg.)

Sample Location OFF01- 112508 Start Time: 1810 End Time: 0705

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #: Tisch TE-1000 Sampler ID #: 1075

Beginning Pressure Drop (Inches H2O): 58 Beginning Timer Reading (hours): 00302.39

Ending Pressure Drop (Inches H2O): Ending Timer Reading (hours): 00306.56 4.17

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #: Tisch TE-1000 Sampler ID #: 1077

Beginning Pressure Drop (Inches H2O): 56 Beginning Timer Reading (hours): 00340.09

Ending Pressure Drop (Inches H2O): Ending Timer Reading (hours): 00350.86 4.17

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #: Tisch TE-1000 Sampler ID #: 1074

Beginning Pressure Drop (Inches H2O): 54 Beginning Timer Reading (hours): 00419.24

Ending Pressure Drop (Inches H2O): Ending Timer Reading (hours): 00423.41 4.17

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 08T6089 Beginning Flow Rate (LPM): 16.7 Battery % 99 / 79

Timer Beginning Time: 0 Ending Flow Rate (LPM): 16.7

Timer Ending Time: 0785 BGI PQ100 Sampler Serial #: 08710

SKC Asbestos Sampler

MCE Filter Lot ID #: Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0 Ending Timer Reading: 0785

Ending Flow Rate (LPM): 2.0 SKC Pump Serial #: 2297

Sample Location OFF02- 112508 Start Time: 1045 End Time: 0725

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #: Tisch TE-1000 Sampler ID #: 1079

Beginning Pressure Drop (Inches H2O): 60 Beginning Timer Reading (hours): 00358.74

Ending Pressure Drop (Inches H2O): 54 Ending Timer Reading (hours): 00363.87 12.13

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #: Tisch TE-1000 Sampler ID #: 1076

Beginning Pressure Drop (Inches H2O): 60 Beginning Timer Reading (hours): 00344.91

Ending Pressure Drop (Inches H2O): 52 Ending Timer Reading (hours): 00359.15 12.24

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #: Tisch TE-1000 Sampler ID #: 1078

Beginning Pressure Drop (Inches H2O): 56 Beginning Timer Reading (hours): 00454.39

Ending Pressure Drop (Inches H2O): 56 Ending Timer Reading (hours): 00460.51 12.12

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 08T6088 Beginning Flow Rate (LPM): 16.7 Battery % 99 / 88

Timer Beginning Time: 0 Ending Flow Rate (LPM): 16.7

Timer Ending Time: 0775 BGI PQ100 Sampler Serial #: 11073

SKC Asbestos Sampler

MCE Filter Lot ID #: Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0 Ending Timer Reading: 0775

Ending Flow Rate (LPM): 2.0 SKC Pump Serial #: 10154

Comments: OFF01 not running on refuel arrival - tripped circuit on generator re started. OFF02 not running in AM when arrive to collect samples. Will call United Rentals for svc.

Signature: [Signature]

TETRA TECH EM INC.
AIR QUALITY SAMPLE DOCUMENTATION SHEET
BMI COMPLEX SITE
SITE INFORMATION

Daily Temperature:

Beginning Ambient Baro. Press. (In. Hg): Ending Ambient Baro. Pressure (In Hg.)

Sample Location OFF01-120208

Start Time: 1815

End Time: 0700

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1075

Beginning Pressure Drop (Inches H2O): 59

Beginning Timer Reading (hours): 00306.56

Ending Pressure Drop (Inches H2O): 54

Ending Timer Reading (hours): 00378.99

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1077

Beginning Pressure Drop (Inches H2O): 64

Beginning Timer Reading (hours): 00350.27

Ending Pressure Drop (Inches H2O): 62

Ending Timer Reading (hours): 00362.69

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1074

Beginning Pressure Drop (Inches H2O): 59

Beginning Timer Reading (hours): 00423.41

Ending Pressure Drop (Inches H2O): 54

Ending Timer Reading (hours): 00435.84

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 08T10694

Beginning Flow Rate (LPM): 16.7

Battery % 99 / -A

Timer Beginning Time: 0

Ending Flow Rate (LPM): 16.7

Timer Ending Time: 0777

BGI PQ100 Sampler Serial #: 11073

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 0777

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 110154

Sample Location OFF02-120208

Start Time: 1845

End Time: 0725

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1079

Beginning Pressure Drop (Inches H2O): 54

Beginning Timer Reading (hours): 00303.87

Ending Pressure Drop (Inches H2O): 48

Ending Timer Reading (hours): 00375.77

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1076

Beginning Pressure Drop (Inches H2O): 50

Beginning Timer Reading (hours): 00357.15

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00369.07

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1078

Beginning Pressure Drop (Inches H2O): 54

Beginning Timer Reading (hours): 00406.51

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00478.42

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 08T10693

Beginning Flow Rate (LPM): 16.7

Battery % 99 / 85

Timer Beginning Time: 0

Ending Flow Rate (LPM): 16.7

Timer Ending Time: 0752

BGI PQ100 Sampler Serial #: 8710

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 0752

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 2997

Comments:

Signature:

TETRA TECH EM INC.
AIR QUALITY SAMPLE DOCUMENTATION SHEET
BMI COMPLEX SITE
SITE INFORMATION

Daily Temperature: Hi 54°F Low 20°F

Beginning Ambient Baro. Press. (in. Hg): 29.4

Ending Ambient Baro. Pressure (in Hg.)

30.4 12-27-2008

Sample Location OFF01- 122708

Start Time: 1745

End Time: 0800

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1075

Beginning Pressure Drop (Inches H2O): 59

Beginning Timer Reading (hours): 00379.16

Ending Pressure Drop (Inches H2O): 59

Ending Timer Reading (hours): 00392.85

13.69

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1077

Beginning Pressure Drop (Inches H2O): 50

Beginning Timer Reading (hours): 00362.87

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00376.55

13.68

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1074

Beginning Pressure Drop (Inches H2O): 54

Beginning Timer Reading (hours): 00436.01

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00449.70

13.69

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 08T6883

Beginning Flow Rate (LPM): 16.7

Battery %

Timer Beginning Time: 0

Ending Flow Rate (LPM): 16.7

Timer Ending Time: 0855 (14.28)

BGI PQ100 Sampler Serial #: 80710

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 0855

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 8850 10154

Sample Location OFF02- 122708

Start Time: 1750

End Time: 0816

Tisch TE-1000/EPA TO 4A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1079

Beginning Pressure Drop (Inches H2O): 54

Beginning Timer Reading (hours): 00375.77

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00389.62

13.80

Tisch TE-1000/EPA TO 9A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1076

Beginning Pressure Drop (Inches H2O): 100

Beginning Timer Reading (hours): 00369.07

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00382.91

13.83

Tisch TE-1000/EPA TO 13A Sampler

Lot/Media ID #:

Tisch TE-1000 Sampler ID #: 1078

Beginning Pressure Drop (Inches H2O): 54

Beginning Timer Reading (hours): 00478.42

Ending Pressure Drop (Inches H2O): 50

Ending Timer Reading (hours): 00492.26

13.81

BGI PQ100 TSP Sampler

47mm Teflon Filter ID #: 08T6884

Beginning Flow Rate (LPM): 16.7

Battery % 79

Timer Beginning Time: 0

Ending Flow Rate (LPM): 16.7

Timer Ending Time: 0816 (14.44)

BGI PQ100 Sampler Serial #: 11673

SKC Asbestos Sampler

MCE Filter Lot ID #:

Beginning Timer Reading: 0

Beginning Flow Rate (LPM): 2.0

Ending Timer Reading: 855 0816

Ending Flow Rate (LPM): 2.0

SKC Pump Serial #: 2991

Comments:

Signature: Beehi Dams