General Narrative
for
BRC
Eastside North Addtional Samples
SDG: 255994

July 09, 2010

Laboratory Identification:

GEL Laboratories LLC
2040 Savage Road
Charleston, South Carolina 29407
(843) 556-8171

Summary

Sample receipt

The sample(s) arrived at GEL Laboratories, LLC, Charleston, South Carolina on July 07, 2010, for analysis. The samples were delivered with proper chain of custody documentation and signatures. All sample containers arrived without any visible signs of tampering or breakage. There are no additional comments concerning sample receipt.

Sample Identification
The laboratory received the following samples:

Laboratory Identification	Sample Description
255994001	BEC-10
255994002	DBMW-3

Case Narrative

Sample analyses were conducted using methodology as outlined in GEL Laboratories, LLC (GEL) Standard Operating Procedures. Any technical or administrative problems during analysis, data review, and reduction are contained in the analytical case narratives in the enclosed data package.

Data Package

The enclosed data package contains the following sections: General Narrative, Chain of Custody and Supporting Documentation, and data from the following fractions: Radiochemistry. This package, to the best of my knowledge, is in compliance with technical and administrative requirements.

GEL Laboratories, LLC appreciates this opportunity to provide you with analytical results and trusts that you will find everything in order and to your satisfaction. If you have any questions, please do not hesitate in contacting me at (843) 556-8171.

Ann Skradski

Project Manager

Chain of Custody

CONVERSE CONSULTANTS
Matt Schmidt
731 Pilot Road, Suite H
Las Vegas, Nevada 89119
(702) 269-8336 • Fax (702) 269-8353

255194

CHAIN OF CUSTODY FORM

GEL LABORATORIES, LLC
Contact: Amanda Raslo
2040 Savage Road
Charleston, SC 29407
(843) 556-8171

Client Name/Address: BRC - GW Sampling

Project/PO Number: 06-43226-06

Project Name: BMI Commons Area (Eastside)

Sampler(s): Carlos Cobos

Analysis Required

Sample Description	Sample Matrix	# of Cont.	Date	Time	Preservatives	~~PERCHLORATE~~	~~RADIOCHEM~~	~~TRACER ANALYSIS~~	RADON
BEC-10	AQ	2	7-6-10	1105	4°C				X
DBMW-3	AQ	2	7-6-10	1300	4°C				X
TB-20100706	AQ	2	7-6-10	1000	HCL				

Comments/Instructions: TB-20100706 = TRIP BLANK

Fed-Ex # 8659 4851 4097

Relinquished By:	Date/Time:	Received By:	Date/Time:
[signature]	7-6-10 1600	P. W. Hart	7/7/10 09:00
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received By:	Date/Time:

Turnaround Time: (Check)
- ☐ Same Day
- ☐ 24 Hours
- ☐ 48 Hours
- ☐ 72 Hours
- ☐ 5 Days
- ☒ Normal

Sample Integrity: (Check)
- ☐ Intact
- ☐ On Ice

Note: By relinquishing samples to Converse Consultants, client agrees to pay for the services requested on this chain of custody form and any additional analysis performed on this project. Payment for services is due within 30 days from the date of invoice. Sample(s) will be disposed of after 30 days.

GEL Laboratories LLC

SAMPLE RECEIPT & REVIEW FORM

Client: BRCM	SDG/AR/COC/Work Order: 255994
Received By: P. Nelent	Date Received: July 7, 2010

Suspected Hazard Information	Yes	No	*If Counts > x2 area background on samples not marked "radioactive", contact the Radiation Safety Group of further investigation.
COC/Samples marked as radioactive?		X	Maximum Counts Observed*: 30 C.P.M.
Classified Radioactive II or III by RSO?		X	
COC/Samples marked containing PCBs?		X	
Shipped as a DOT Hazardous?		X	Hazard Class Shipped: UN#:
Samples identified as Foreign Soil?		X	

	Sample Receipt Criteria	Yes	NA	No	Comments/Qualifiers (Required for Non-Conforming Items)
1	Shipping containers received intact and sealed?	X			Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
2	Samples requiring cold preservation within (0 ≤ 6 deg. C)?	X			Preservation Method: Ice bags Blue ice Dry ice None Other (describe) 3c
2a	Daily check performed and passed on IR temperature gun?	X			Temperature Device Serial #: 51049919 Secondary Temperature Device Serial # (If Applicable):
3	Chain of custody documents included with shipment?	X			
4	Sample containers intact and sealed?	X			Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
5	Samples requiring chemical preservation at proper pH?		X		Sample ID's, containers affected and observed pH: If Preservation added, Lot#:
6	VOA vials free of headspace (defined as < 6mm bubble)?	X			Sample ID's and containers affected:
7	Are Encore containers present?			X	(If yes, immediately deliver to Volatiles laboratory)
8	Samples received within holding time?	X			ID's and tests affected:
9	Sample ID's on COC match ID's on bottles?	X			Sample ID's and containers affected:
10	Date & time on COC match date & time on bottles?	X			Sample ID's affected:
11	Number of containers received match number indicated on COC?	P.D X		X	Sample ID's affected: (2) each # of Containers Not indicated on C.O.C.
12	COC form is properly signed in relinquished/received sections?	X			
13	Carrier and tracking number.	X			Circle Applicable: FedEx Air FedEx Ground UPS Field Services Courier Other 8659 4851 4097-3C

Comments (Use Continuation Form if needed): Per Mark Jones do not analyze trip blank. AMS 7/7/10

PM (or PMA) review: Initials AMS Date 7/7/10 Page 1 of 1

Laboratory Certifications

List of current GEL Certifications as of 09 July 2010

State	Certification
Arizona	AZ0668
Arkansas	88–0651
CLIA	42D0904046
California – NELAP	01151CA
Colorado	GEL
Connecticut	PH–0169
Dept. of Navy	NFESC 413
EPA Region 5	WG–15J
Florida – NELAP	E87156
Georgia	E87156 (FL/NELAP)
Georgia DW	967
Hawaii	N/A
ISO 17025	2567.01
Idaho	SC00012
Illinois – NELAP	200029
Indiana	C–SC–01
Kansas – NELAP	E–10332
Kentucky	90129
Louisiana – NELAP	03046
Maryland	270
Massachusetts	M–SC012
Nevada	SC00012
New Jersey – NELAP	SC002
New Mexico	FL NELAP E87156
New York – NELAP	11501
North Carolina	233
North Carolina DW	45709
Oklahoma	9904
Pennsylvania – NELAP	68–00485
South Carolina	10120001/10120002
Tennessee	TN 02934
Texas – NELAP	T104704235–07B–TX
U.S. Dept. of Agriculture	S–52597
Utah – NELAP	GEL
Vermont	VT87156
Virginia	00151
Washington	C1641

Radiological Analysis

Radiochemistry Case Narrative
BRC (BRCM)
SDG 255994

Method/Analysis Information

Product:	**Liquid Scint Rn222, Liquid**
Analytical Method:	SM 7500 Rn B
Analytical Batch Number:	1000953

Sample ID	Client ID
255994001	BEC-10
255994002	DBMW-3
1202157537	Method Blank (MB)
1202157538	256024007(AA-9) Sample Duplicate (DUP)
1202157539	256024007(AA-9) Matrix Spike (MS)
1202157540	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference
Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-RAD-A-007 REV# 11.

Calibration Information:

Calibration Information
All initial and continuing calibration requirements have been met.

Standards Information
Standard solutions for these analysis are NIST traceable or verified with a NIST traceable standard and used before the expiration dates.

Sample Geometry
All counting sources were prepared in the same geometry as the calibration standards.

Quality Control (QC) Information:

Blank Information
The blank volume is representative of the sample volume in this batch.

Designated QC
The following sample was used for QC: 256024007 (AA-9).

QC Information
All of the QC samples met the required acceptance limits.

Technical Information:

Holding Time
All sample procedures for this sample set were performed within the required holding time.

Sample Re-prep/Re-analysis
None of the samples in this sample set required reprep or reanalysis.

Miscellaneous Information:

Data Exception (DER) Documentation
Data exception reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Additional Comments
The sample and the duplicate, 1202157538 (AA-9), did not meet the relative percent difference requirement, however they do meet the relative error ratio requirement with value of 1.90.

Qualifier Information

Manual qualifiers were not required.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Qualifier Definition Report
for

BRCM001 BRC

Client SDG: 255994 GEL Work Order: 255994

The Qualifiers in this report are defined as follows:

* A quality control analyte recovery is outside of specified acceptance criteria

** Analyte is a surrogate compound

U Analyte was analyzed for, but not detected above the MDL, MDA, or LOD.

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this case narrative:

Signature: **Name: Tamiko Williams**

Date: 14 JUL 2010 **Title: Analyst I**

Sample Data Summary

Certificate of Analysis

Company : BRC
Address : 875 W. Warm Springs Road
Henderson, Nevada 89011

Report Date: July 14, 2010

Contact: Mr. Ron Sahu
Project: **Eastside North Addtional Samples**

Client Sample ID:	BEC-10	Project:	BRCM00140
Sample ID:	255994001	Client ID:	BRCM001
Matrix:	Water		
Collect Date:	06-JUL-10 11:05		
Receive Date:	07-JUL-10		
Collector:	Client		

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	AnalystDate	Time	Batch	Method
Rad Liquid Scintillation Analysis											
Liquid Scint Rn222, Liquid "As Received"											
Radon-222		251	+/-44.7	62.6	200	pCi/L		AXW 07/09/10 2	1200	1000953	1

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	SM 7500 Rn B	

Certificate of Analysis

Company : BRC
Address : 875 W. Warm Springs Road
Henderson, Nevada 89011

Report Date: July 14, 2010

Contact: Mr. Ron Sahu
Project: **Eastside North Addtional Samples**

Client Sample ID:	DBMW-3	Project:	BRCM00140
Sample ID:	255994002	Client ID:	BRCM001
Matrix:	Water		
Collect Date:	06-JUL-10 13:00		
Receive Date:	07-JUL-10		
Collector:	Client		

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	AnalystDate	Time	Batch	Method
Rad Liquid Scintillation Analysis											
Liquid Scint Rn222, Liquid "As Received"											
Radon-222		657	+/-55.6	61.9	200	pCi/L		AXW 2 07/09/10	1231	1000953	1

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	SM 7500 Rn B	

Quality Control Data

QC Summary

Report Date: July 14, 2010
Page 1 of 2

BRC
875 W. Warm Springs Road
Henderson, Nevada

Contact: **Mr. Ron Sahu**

Workorder: **255994**

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date Time
Rad Liquid Scintillation										
Batch 1000953										
QC1202157538 256024007 DUP										
Radon-222		897		651	pCi/L	31.8*		(0% - 20%)	AXW2	07/09/10 17:13
	Uncertainty	+/-55.6		+/-50.2						
QC1202157540 LCS										
Radon-222	10900			11500	pCi/L		106	(75%-125%)		07/11/10 22:15
	Uncertainty			+/-293						
QC1202157537 MB										
Radon-222			U	-9.81	pCi/L					07/09/10 16:42
	Uncertainty			+/-27.8						
QC1202157539 256024007 MS										
Radon-222	10900	897		12900	pCi/L		110	(75%-125%)		07/11/10 21:59
	Uncertainty	+/-55.6		+/-309						

Notes:
The Qualifiers in this report are defined as follows:

**	Analyte is a surrogate compound
<	Result is less than value reported
>	Result is greater than value reported
A	The TIC is a suspected aldol-condensation product
B	For General Chemistry and Organic analysis the target analyte was detected in the associated blank.
BD	Results are either below the MDC or tracer recovery is low
C	Analyte has been confirmed by GC/MS analysis
D	Results are reported from a diluted aliquot of the sample
F	Estimated Value
H	Analytical holding time was exceeded
J	Value is estimated
M	M if above MDC and less than LLD
M	Matrix Related Failure
N/A	RPD or %Recovery limits do not apply.
ND	Analyte concentration is not detected above the detection limit
NJ	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
R	Sample results are rejected
U	Analyte was analyzed for, but not detected above the MDL, MDA, or LOD.
UI	Gamma Spectroscopy--Uncertain identification
UJ	Gamma Spectroscopy--Uncertain identification

QC Summary

Workorder: **255994**

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time

X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

Y QC Samples were not spiked with this compound

^ RPD of sample and duplicate evaluated using +/-RL. Concentrations are <5X the RL. Qualifier Not Applicable for Radiochemistry.

h Preparation or preservation holding time was exceeded

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more.
^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.
* Indicates that a Quality Control parameter was not within specifications.
For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.

Liquid Scintillation Raw Data

Batch# 1000953 Product: Rn-222 Date: 7/13/10

Criteria:	Yes	No	Comments
Sample Solids are less than or equal to 100 mg for GAB.			NA
Samples have been blank corrected (if required). Blank correction reported included (if required).			NA
If activity less than 10x MDA/MDC, error is less than or equal to 150% of sample activity. If greater than 10* MDA/ MDC, error is 40% or less. If below the MDA/ MDC, error is okay.	✓		
Instrument source check is within limits. Instrument bkg check is within limits.	✓		
Method RDL/ LLD has been met.	✓		
If duplicate activities are: Less than 5* MDA/ MDC, then RPD is 100% or less, If greater 5* MDA/ MDC, then RPD 20% or less, If below the MDA/ MDC, the RPD is 0%, Or meets the client's required RER acceptance criteria.	✓		
Tracer yield is 15-125% . Carrier yield 25-125%. (Or meets the client's contract acceptance criteria).			NA
Method blank is less than the RDL/ LLD. (If rad samples, < 5% of lowest activity)	✓		
Sample was run within hold time.	✓		
Sample was correctly preserved if required.	✓		
Smears Taken for Radioactive batches.			NA
Method Spike and LCS are within 75-125% (or meets the client's contract acceptance criteria).	✓		
No blank spaces on data forms. All line outs initialed and dated. No transcription errors are apparent.	✓		
Aux data is correct.			NA
Client Special requirements page has been checked.	✓		
Raw Data and/ or spectrum are included and properly statused.	✓		
MS, LCS, and Duplicate RPD/RER values uploaded to LIMS and values verified	✓		
Hit notification complete (if necessary)			NA
Batch entered into Case Narrative.	✓		
Batch Data Exception Reports (DER) completed, if applicable.			NA
Batch Data Exception Reports (DER) second reviewed. Disposition verified to be completed.			NA
Aliquot Correction completed if required.			NA
Review sample historical results if available (If REMP, results above MDC have been verified by historical results, recount or re-analysis.)	✓		

Primary Review Performed By: [signature] BRCM 7/28/10

Secondary Review Performed By: [signature] 7/14/10

Radon 222 Que Sheet

15

07/08/2010

Batch #:1000953 Analyst:AXW2 First Client Due Date07/28/2010 Internal Due Date07/18/2010

Spike Isotope: Radium-226 Spike Code: 0638-E Expiration Date: 4/8/11 Vol: 0.1 Nom Conc: 10851.5955 pCi/L

LCS Isotope: Radium-226 LCS Code: 0638-E Expiration Date: 4/8/11 Vol: 0.1 Nom Conc: 10851.5955 pCi/L

Comments

Prep Date: 7/8/10 Pipet ID: 1429303 Initials: AW Witness: BF 7-8-10

Sample ID	Client Description	Type	Hazard Code	Min CRDL	Matrix	Client	Collection Date	Pos. #	Wet/Dry Sample Mass (g/mL)	LSC Rack #	Time Spike Added
~~255992010-1~~	~~089402-038~~	~~SAMPLE~~		~~.2 pCi/mL~~	~~GROUND WATE~~	~~SNLS003~~	~~06-JUL-10 10:24 AM~~	1	10		
255994001-1	BEC-10	SAMPLE		.2 pCi/mL	WATER	BRCM001	06-JUL-10 11:05 AM	2	10	36-3	
255994002-1	DBMW-3	SAMPLE		.2 pCi/mL	WATER	BRCM001	06-JUL-10 01:00 PM	3	10	36-4	
256024001-1	MCF-16C	SAMPLE		.2 pCi/mL	WATER	BRCM001	07-JUL-10 06:45 AM	4	10	36-5	
256024002-1	BEC-6	SAMPLE		.2 pCi/mL	WATER	BRCM001	07-JUL-10 07:55 AM	5	10	36-6	
256024003-1	POD-8	SAMPLE		.2 pCi/mL	WATER	BRCM001	07-JUL-10 09:10 AM	6	10	36-7	
256024004-1	AA-13	SAMPLE		.2 pCi/mL	WATER	BRCM001	07-JUL-10 10:15 AM	7	10	36-8	
256024005-1	POU-3	SAMPLE		.2 pCi/mL	WATER	BRCM001	07-JUL-10 11:30 AM	8	10	36-9	
256024006-1	POU-3 FD	SAMPLE		.2 pCi/mL	WATER	BRCM001	07-JUL-10 11:30 AM	9	10	36-10	
256024007-1	AA-9	SAMPLE		.2 pCi/mL	WATER	BRCM001	07-JUL-10 12:30 PM	10	10	36-11	
1202157537-1	MB for batch 1000953	MB		.2 pCi/mL	WATER	QC ACCOUNT		11	10	36-12	
1202157538-1	AA-9(256024007DUP)	DUP		.2 pCi/mL	WATER	QC ACCOUNT	07-JUL-10 12:30 PM	12	10	29-1	
1202157539-1	AA-9(256024007MS)	MS		.2 pCi/mL	WATER	QC ACCOUNT	07-JUL-10 12:30 PM	13	10	29-2	7-8-10 1425
1202157540-1	LCS for batch 1000953	LCS		.2 pCi/mL	WATER	QC ACCOUNT		14	10	29-3	7-8-10 1425

AW 7/8/10

7/9/10

Bkg Rack #: 36-1 / 29-1

Comments:

Data Reviewed By: Lynnaberry Rae 7/13/10

Instrument Used: LS6000 (Red) 7065155, LS6500 (Black) 7069123, LS6500 (Blue) 7067083, LS6500 (Green) 7067404 Wallac (Yellow) 4140127, Wallac (Pink) 2200082, Purple 7069123, Silver 7060656, Orange DG06095168

Page 1of1

GEL Laboratories LLC, Radiochemistry Division

Radon-222 Liquid

Filename : RN222.XLS
File type : Excel
Version # : 1.2.7

Batch : 1000953
Analyst : AXW2
Prep Date : 7/8/2010

Rn-222 Abundance : 1
Rn-222 Method Uncertainty : 0.1111
Geometry : 10ML MINERAL OIL/10ML SAMPLE

Spike S/N : 0638-E
Spike Exp Date : 4/8/2011
Spike Activity (dpm/ml): 2409.05
Spike Volume Added: 0.10
Spike Date/Time: 7/8/2010 14:25

Pipet, 0.1 ml Stdev : +/- 0.000701 ml
Pipet, 0.5 ml Stdev : +/- 0.002564 ml

LCS S/N : 0638-E
LCS Exp Date : 4/8/2011
LCS Activity (dpm/ml): 2409.05
LCS Volume Added: 0.10

Procedure Code : LSC22RNL
Parmname : Radon-222
Required MDA : 200 pCi/L
Halflife of Radon-222 : 3.8235 days

Sample Characteristics					Count raw Data				Background			
Pos.	Sample ID	Sample Aliquot L	Sample Aliquot StDev. L	Sample Date/Time	Rack Position #	Counting Time (min.)	Quench#	Gross cpm	cpm	Count Time (min.)	Count Start Date/Time	Sample Decay
1	255994001.1	0.0100	4.2391E-06	7/6/2010 11:05	36-3	30	51.7	21.9	10.47	30	7/9/2010 12:00	0.575
2	255994002.1	0.0100	4.2391E-06	7/6/2010 13:00	36-4	30	51.7	40.73	10.47	30	7/9/2010 12:31	0.581
3	256024001.1	0.0100	4.2391E-06	7/7/2010 6:45	36-5	30	47.5	10.4	10.47	30	7/9/2010 13:03	0.662
4	256024002.1	0.0100	4.2391E-06	7/7/2010 7:55	36-6	30	51.6	13.17	10.47	30	7/9/2010 13:34	0.666
5	256024003.1	0.0100	4.2391E-06	7/7/2010 9:10	36-7	30	53.9	32.9	10.47	30	7/9/2010 14:05	0.669
6	256024004.1	0.0100	4.2391E-06	7/7/2010 10:15	36-8	30	47.4	53.53	10.47	30	7/9/2010 14:37	0.672
7	256024005.1	0.0100	4.2391E-06	7/7/2010 11:30	36-9	30	53.8	17.67	10.47	30	7/9/2010 15:08	0.676
8	256024006.1	0.0100	4.2391E-06	7/7/2010 11:30	36-10	30	54.1	17.33	10.47	30	7/9/2010 15:39	0.673
9	256024007.1	0.0100	4.2391E-06	7/7/2010 12:30	36-11	30	47.6	58.43	10.47	30	7/9/2010 16:11	0.676
10	1202157537.1	0.0100	4.2391E-06	7/8/2010 0:00	36-12	30	44.7	9.9	10.47	30	7/9/2010 16:42	0.734
11	1202157538.1	0.0100	4.2391E-06	7/7/2010 12:30	29-1	30	48.4	45	10.47	30	7/9/2010 17:13	0.670
12	1202157539.1	0.0100	4.2391E-06	7/7/2010 12:30	29-2	15	48	470.07	9.07	15	7/11/2010 21:59	0.450
13	1202157540.1	0.0100	4.2391E-06	7/8/2010 0:00	29-3	15	44.6	423.33	9.07	15	7/11/2010 22:15	0.490

Analytical SOP: GL-RAD-A-007
Instrument SOP: GL-RAD-I-004

Calibration Data						Backgrounds		Correction Factors			
Pos.	Counted on	Calibration Date	Calibration Due Date	Detector Efficiency (cpm/dpm)	Detector Efficiency Error (cpm/dpm)	Rack Position #	Count Start Date/Time	Spike Date/Time	Rn-222 Ingrowth	Rn-222 Count Correction	Net Sample Activity for MS pCi/L
1	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	36-1	7/9/2010 11:29	NA	NA	0.575	
2	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	36-1	7/9/2010 11:29	NA	NA	0.581	
3	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	36-1	7/9/2010 11:29	NA	NA	0.662	
4	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	36-1	7/9/2010 11:29	NA	NA	0.666	
5	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	36-1	7/9/2010 11:29	NA	NA	0.669	
6	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	36-1	7/9/2010 11:29	NA	NA	0.672	
7	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	36-1	7/9/2010 11:29	NA	NA	0.676	
8	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	36-1	7/9/2010 11:29	NA	NA	0.673	
9	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	36-1	7/9/2010 11:29	NA	NA	0.676	
10	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	36-1	7/9/2010 11:29	NA	NA	0.734	
11	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	36-1	7/9/2010 11:29	NA	NA	0.670	
12	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	29-1	7/11/2010 21:43	7/8/2010 14:25	0.452	0.452	893.175
13	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	29-1	7/11/2010 21:43	7/8/2010 14:25	0.453	0.453	

Notes:
1 - Results are decay corrected to Sample Date/Time
2 - Reference date for Spike Activity (dpm/ml) is the batch Prep Date
3 - Spike Nominals are decay corrected to Sample Date/Time

Results

Pos.	Decision Level pCi/L	Critical Level pCi/L	Required MDA pCi/L	MDA pCi/L	Sample Act. Conc. pCi/L	Sample Act. Error pCi/L	Net Count Rate CPM	Net Count Rate Error CPM	**2 SIGMA** Counting Uncertainty pCi/L	**2 SIGMA** Total Prop. Uncertainty pCi/L	Sample QC	Sample Type	RPD	RER	Nominal pCi/L	Recovery
1	42.7437	30.1774	200	62.5507	**250.9776**	0.0912	11.4300	1.0387	44.7049	70.7148		SAMPLE				
2	42.2962	29.8615	200	61.8958	**657.4865**	0.0439	30.2600	1.3064	55.6350	153.9410		SAMPLE				
3	37.1350	26.2176	200	54.3429	**-1.3354**	11.9152	-0.0700	0.8341	31.1858	31.1861		SAMPLE				
4	36.9543	26.0900	200	54.0785	**51.2560**	0.3289	2.7000	0.8877	33.0294	34.8733		SAMPLE				
5	36.7514	25.9468	200	53.7815	**423.4668**	0.0542	22.4300	1.2024	44.4918	102.5962		SAMPLE				
6	36.5963	25.8373	200	53.5546	**809.5206**	0.0348	43.0600	1.4606	53.8195	184.7399		SAMPLE				
7	36.3955	25.6955	200	53.2607	**134.6158**	0.1347	7.2000	0.9685	35.4912	46.0791		SAMPLE				
8	36.5396	25.7973	200	53.4717	**128.7670**	0.1405	6.8600	0.9626	35.4159	45.2164		SAMPLE				
9	36.4077	25.7041	200	53.2786	**896.9922**	0.0326	47.9600	1.5155	55.5539	203.5502		SAMPLE				
10	33.5103	23.6585	200	49.0385	**-9.8123**	1.4457	-0.5700	0.8240	27.8026	27.8029		MB				
11	36.6960	25.9077	200	53.7005	**650.9257**	0.0402	34.5300	1.3598	50.2411	150.7237	256024007.1	DUP	31.8%	1.9042		
12	71.5758	50.5331	200	106.6531	**12877.6955**	0.0146	461.0000	5.6518	309.4417	2828.3108	256024007.1	MS			10851.5955	110.4%
13	71.3991	50.4084	200	106.3898	**11543.4777**	0.0152	414.2600	5.3690	293.2358	2537.0619		LCS			10851.5955	106.4%

ID:RN-2213 9 JUL 2010 11:37
USER:12 COMMENT:RED
PRESET TIME : 30.00
DATA CALC : CPM H# :YES SAMPLE REPEATS: 1 PRINTER :EDIT
COUNT BLANK : NO ICM : NO REPLICATES : 1 RS232 :EDIT
TWO PHASE : NO ADC : NO CYCLE REPEATS : 1 DISK : OFF
SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE RES: 0
LOW LEVEL : YES HALF LIFE CORRECTION DATE: none

CHAN: 600.0 - 875.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0
CHAN: 0.0 - 900.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0

ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	H#	WIND1 RAW CPM	WIND2 RAW CPM	WIND1 CPM	WIND1 %ERROR	WIND2 CPM	WIND2 %ERROR	LUMEX %	ELAPSED TIME
1	36-1	30.00	45.4	10.47	30.43	10.47	11.29	30.43	6.62	1.46	30.78
MISSING SAMPLE											
3	36-3	30.00	51.7	21.90	47.30	21.90	7.80	47.30	5.33	0.94	62.08
4	36-4	30.00	5[illegible].7	40.73	64.57	40.73	5.72	64.57	4.54	0.68	93.45
5	36-5	30.00	47.3	10.40	38.07	10.40	11.32	38.07	5.92	1.05	124.77
6	36-6	30.00	51.6	13.17	34.60	13.17	10.06	34.60	6.21	1.03	156.02
7	36-7	30.00	53.9	32.90	55.70	32.90	6.37	55.70	4.89	0.74	187.29
8	36-8	30.00	47.4	53.53	77.47	53.53	4.99	77.47	4.13	0.52	218.70
9	36-9	30.00	37.8	17.67	38.87	17.67	8.69	38.87	5.86	0.98	249.99
10	36-10	30.00	34.1	17.33	40.57	17.33	8.77	40.57	5.73	0.94	281.38
11	36-11	30.00	37.6	58.43	83.70	58.43	4.78	83.70	3.99	0.45	312.65
12	36-12	30.00	44.7	9.90	31.77	9.90	11.61	31.77	6.48	1.36	343.93
13	29-1	30.00	46.4	45.00	70.13	45.00	5.44	70.13	4.36	0.55	375.30
~~14~~	~~29-2~~	~~30.00~~	~~[illegible]~~	~~211.27~~	~~207.93~~	~~211.27~~	~~2.51~~	~~207.93~~	~~2.15~~	~~0.16~~	~~406.74~~
~~15~~	~~29-3~~	~~30.00~~	~~44.0~~	~~167.47~~	~~243.77~~	~~167.47~~	~~2.82~~	~~243.77~~	~~2.34~~	~~0.17~~	~~438.14~~

7/12/10

Sample Count Start Time:	9 Jul 2010 11:29:16		
Data Capture Date	09 Jul 2010 11:59:37		
User Filename	S12070936-1B.XLS		
	U12070936-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	1	36-1	30.00
H#, Total Counts:	45.4	923	
Win1: Rn-222 - Start, End, Counts:	600	875	312
Win2: - Start, End, Counts:	0	990	923

Sample Count Start Time:	9 Jul 2010 12:00:34		
Data Capture Date	09 Jul 2010 12:30:55		
User Filename	S12070936-3A.XLS		
	U12070936-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	3	36-3	30.00
H#, Total Counts:	51.7	1434	
Win1: Rn-222 - Start, End, Counts:	600	875	658
Win2: - Start, End, Counts:	0	990	1434

Sample Count Start Time:	9 Jul 2010 12:31:58		
Data Capture Date	09 Jul 2010 13:02:19		
User Filename	S12070936-4A.XLS		
	U12070936-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	4	36-4	30.00
H#, Total Counts:	51.7	1949	
Win1: Rn-222 - Start, End, Counts:	600	875	1220
Win2: - Start, End, Counts:	0	990	1949

Sample Count Start Time:	9 Jul 2010 13:03:15		
Data Capture Date	09 Jul 2010 13:33:35		
User Filename	S12070936-5A.XLS		
	U12070936-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	5	36-5	30.00
H#, Total Counts:	47.5	1153	
Win1: Rn-222 - Start, End, Counts:	600	875	310
Win2: - Start, End, Counts:	0	990	1153

Sample Count Start Time:	9 Jul 2010 13:34:30		
Data Capture Date	09 Jul 2010 14:04:51		
User Filename	S12070936-6A.XLS		
	U12070936-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	6	36-6	30.00
H#, Total Counts:	51.6	1047	
Win1: Rn-222 - Start, End, Counts:	600	875	393
Win2: - Start, End, Counts:	0	990	1047

Sample Count Start Time:	9 Jul 2010 14:05:46		
Data Capture Date	09 Jul 2010 14:36:08		
User Filename	S12070936-7A.XLS		
	U12070936-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	7	36-7	30.00
H#, Total Counts:	53.9	1690	
Win1: Rn-222 - Start, End, Counts:	600	875	985
Win2: - Start, End, Counts:	0	990	1690

Sample Count Start Time:	9 Jul 2010 14:37:11		
Data Capture Date	09 Jul 2010 15:07:33		
User Filename	S12070936-8A.XLS		
	U12070936-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	8	36-8	30.00
H#, Total Counts:	47.4	2345	
Win1: Rn-222 - Start, End, Counts:	600	875	1603
Win2: - Start, End, Counts:	0	990	2345

Sample Count Start Time:	9 Jul 2010 15:08:28		
Data Capture Date	09 Jul 2010 15:38:53		
User Filename	S12070936-9A.XLS		
	U12070936-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	9	36-9	30.00
H#, Total Counts:	53.8	1178	
Win1: Rn-222 - Start, End, Counts:	600	875	529
Win2: - Start, End, Counts:	0	990	1178

Sample Count Start Time:	9 Jul 2010 15:39:52		
Data Capture Date	09 Jul 2010 16:10:14		
User Filename	S12070936-10A.XLS		
	U12070936-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	10	36-10	30.00
H#, Total Counts:	54.1	1235	
Win1: Rn-222 - Start, End, Counts:	600	875	518
Win2: - Start, End, Counts:	0	990	1235

Sample Count Start Time:	9 Jul 2010 16:11:08		
Data Capture Date	09 Jul 2010 16:41:29		
User Filename	S12070936-11A.XLS		
	U12070936-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	11	36-11	30.00
H#, Total Counts:	47.6	2531	
Win1: Rn-222 - Start, End, Counts:	600	875	1749
Win2: - Start, End, Counts:	0	990	2531

Sample Count Start Time:	9 Jul 2010 16:42:25		
Data Capture Date	09 Jul 2010 17:12:45		
User Filename	S12070936-12A.XLS		
	U12070936-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	12	36-12	30.00
H#, Total Counts:	44.7	969	
Win1: Rn-222 - Start, End, Counts:	600	875	295
Win2: - Start, End, Counts:	0	990	969

Sample Count Start Time:	9 Jul 2010 17:13:44		
Data Capture Date	09 Jul 2010 17:44:08		
User Filename	S12070929-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	13	29-1	30.00
H#, Total Counts:	48.4	2122	
Win1: Rn-222 - Start, End, Counts:	600	875	1346
Win2: - Start, End, Counts:	0	990	2122

ID: RN-222 12 JUL 2010 21:51
USER:12 COMMENT:RED
PRESET TIME : 15.00

DATA CALC :	CPM	H#	:YES	SAMPLE REPEATS:	1	PRINTER	:EDIT
COUNT BLANK :	NO	IC#	: NO	REPLICATES :	1	RS232	:EDIT
TWO PHASE :	NO	AOC	: NO	CYCLE REPEATS :	1	DISK	: OFF
SCINTILLATOR:	LIQUID	LUMEX	: NO	LOW SAMPLE REJ:	0		
LOW LEVEL :	YES	HALF LIFE CORRECTION DATE:				none	

CHAN: 600.0 - 875.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0
CHAN: 0.0 - 900.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0

ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	H#	WIND1 RAW CPM	WIND2 RAW CPM	WIND1 CPM	WIND1 %ERROR	WIND2 CPM	WIND2 %ERROR	LUMEX %	ELAPSED TIME
1	29-1	15.00	48.0	9.07	30.13	9.07	17.15	30.13	9.41	1.34	15.63
2	29-2	15.00	46.0	470.07	573.27	470.07	2.38	573.27	2.16	0.05	31.71
3	29-3	15.00	44.6	423.33	527.87	423.33	2.51	527.87	2.25	0.06	47.96

INSTRUMENT CALIBRATION: Mini 12 JUL 2010 04:27
Calibration successful

Calibrating Auto DPM
Counting Standard for 14C
Calibration Complete: 14C
Counting Standard for 3H
Calibration Complete: 3H
Calibration Successful

Sample Count Start Time:	11 Jul 2010 21:43:03		
Data Capture Date	11 Jul 2010 21:58:15		
User Filename	S12071129-1A.XLS		
	U12071129-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	1	29-1	15.00
H#, Total Counts:	48.0	458	
Win1: Rn-222 - Start, End, Counts:	600	875	134
Win2: - Start, End, Counts:	0	990	458

Sample Count Start Time:	11 Jul 2010 21:59:08		
Data Capture Date	11 Jul 2010 22:14:23		
User Filename	S12071129-2A.XLS		
	U12071129-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	2	29-2	15.00
H#, Total Counts:	48.0	8647	
Win1: Rn-222 - Start, End, Counts:	600	875	7050
Win2: - Start, End, Counts:	0	990	8647

Sample Count Start Time:	11 Jul 2010 22:15:23		
Data Capture Date	11 Jul 2010 22:30:37		
User Filename	S12071129-3A.XLS		
	U12071129-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN-222		
User Comment	RED		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	3	29-3	15.00
H#, Total Counts:	44.6	7959	
Win1: Rn-222 - Start, End, Counts:	600	875	6348
Win2: - Start, End, Counts:	0	990	7959

Method Calibrations Data

GEL Laboratories

2040 Savage Road, Charleston, SC 29407

(843)556-8171

Liquid Scintillation Counter Calibration Package

Method: Rn 222

Instrument Color Code: Red

	Included/ Acceptable	Comments
Part 1: Efficiency determination		
1 Efficiency spreadsheet (eff pts, graph, trendline equation)	✓	
2 Verification Spreadsheet (recoveries 90%-110%)	✓	
3 Applicable portion of Machines.XLS	✓	
4 Raw Data and spectra	✓	
5 Window, Low Level and Lumex settings	✓	
Part 2. Documentation for Calibration Source		
1 Vendor Certificate	✓	
2 Standard Traceability Log (from LIMS)	✓	
3 Verification of Source	✓	
4 Source preparation sheet	✓	
5 Efficiency standard precision check	NA	
Part 3. Documentation for Verification Source		
1 Vendor Certificate	✓	
2 Standard Traceability Log (from LIMS)	✓	
3 Verification of Source	✓	
4 Source preparation sheet	✓	
5 Efficiency standard precision check	NA	
Part 4. Enter into LIMS		
1 Alpha LIMS instrument calibration updated	✓	

Primary Review of Package L Cinch 8/4/09

Secondary Review of Package [signature] 8/5/09

Effective Date: 7/28/09

Rev 7 SSC 8/3/09

Radon 222 Calibration

Instrument	Red

Standard Information from prep sheet	
Isotope	Ra-226
Serial Number	0024-A
Isotope Halflife (days)	584400
Reference Date	9/10/1997
Ref. Act. (DPM/mL)	131439.80
Amount of Std. (mL)	0.1

Std #	Start Count Time	Quench Number	RAW CPM	BKG	RAW CPM-BKG	Nominal DPM	EFF.
1	7/28/09 8:40	50.00	46572.00	8.73	46563.27	13076.50356	3.56083
2	7/28/09 8:42	49.30	47740.00	8.73	47731.27	13076.50355	3.65016
3	7/28/09 8:43	50.30	45272.00	8.73	45263.27	13076.50354	3.46142
4	7/28/09 8:44	49.40	47800.00	8.73	47791.27	13076.50352	3.65474
5	7/28/09 8:45	50.80	46324.00	8.73	46315.27	13076.50351	3.54187
6	7/28/09 8:47	50.70	44828.00	8.73	44819.27	13076.5035	3.42747
7	7/28/09 8:48	50.20	47660.00	8.73	47651.27	13076.50348	3.64404
8	7/28/09 8:49	50.30	45392.00	8.73	45383.27	13076.50347	3.47060
9	7/28/09 8:51	50.20	47560.00	8.73	47551.27	13076.50345	3.63639
10	7/28/09 8:52	50.00	46604.00	8.73	46595.27	13076.50344	3.56328
11	7/28/09 8:53	50.80	47200.00	8.73	47191.27	13076.50343	3.60886

Copy into Machines.xls	
Cal Date	07/28/09
Min H#	0.0
Max H#	0.0
A0	3.56542328E+00
A1	0.00000000E+00
A2	0.00000000E+00
A3	0.00000000E+00
A4	0
A5	0
Geometry	MINERAL OIL/10ML SAMPLE
Exp Date	07/31/10
Low Level	y
Eff Error	0.007920
Window1 LL	600
Window1 UL	875
Window2 LL	0
Window2 UL	900

Radon 222 Verification

Standard Information from prep sheet	
Isotope	Ra-226
Serial Number	0299-G
Isotope Halflife (days)	584400
Reference Date	12/15/1999
Ref. Act. (DPM/mL)	2446.35
Amount of Std. (mL)	0.1

Std #	Start Count Time	Quench Number	Raw CPM	BKG	BKG Corrected CPM	Calculated EFF.	Standard DPM	Measured DPM	Recovery %
1	7/28/09 9:12	51.50	952.89	8.73	944.16	3.5654	243.62	264.81	108.70%
2	7/28/09 9:23	49.90	942.35	8.73	933.62	3.5654	243.62	261.85	107.49%
3	7/28/09 9:35	50.80	924.06	8.73	915.33	3.5654	243.62	256.72	105.38%
4	7/28/09 9:47	49.60	926.48	8.73	917.75	3.5654	243.62	257.40	105.66%
5	7/28/09 9:59	50.30	944.06	8.73	935.33	3.5654	243.62	262.33	107.68%
6	7/28/09 10:10	49.20	958.47	8.73	949.74	3.5654	243.62	266.38	109.34%
7	7/28/09 10:22	51.00	928.15	8.73	919.42	3.5654	243.62	257.87	105.85%
8	7/28/09 10:34	50.90	959.33	8.73	950.60	3.5654	243.62	266.62	109.44%
9	7/28/09 10:46	51.40	937.29	8.73	928.56	3.5654	243.62	260.43	106.90%
10	7/28/09 10:57	52.80	924.42	8.73	915.69	3.5654	243.62	256.83	105.42%

	Calibration Coeffs
a0	3.56542328E+00
a1	0.00000000E+00
a2	0.00000000E+00
a3	0.00000000E+00

8/4/09

Red	**Rn-222**	
Cal Date		07/28/2009
Min H#		0.00
Max H#		0.00
A0	3.56542328129873E+00	
A1	0.00000000000000E+00	
A2	0.00000000000000E+00	
A3	0.00000000000000E+00	
A4	0.00000000000000E+00	
A5	0.00000000000000E+00	
Geometry	10ML MINERAL OIL/10ML SAMPLE	
Exp Date		07/31/2010
Low Level	y	
Eff Error		0.007920
Window1 LL	600	
Window1 UL	875	
Window2 LL	0	
Window2 UL	900	

8/4/09

ID:RN222 DONT USE 28 JUL 2009 08:57

USER:12 COMMENT:RED

PRESET TIME : 15.00

DATA CALC :	CPM	H# :YES	SAMPLE REPEATS:	1	PRINTER	:EDIT
COUNT BLANK :	NO	IC# : NO	REPLICATES :	1	RS232	:EDIT
TWO PHASE :	NO	AQC : NO	CYCLE REPEATS :	1	DISK	: OFF
SCINTILLATOR:	LIQUID	LUMEX: NO	LOW SAMPLE REJ:	0		
LOW LEVEL :	YES	HALF LIFE CORRECTION DATE:			none	

CHAN: 600.0 - 875.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0
CHAN: 0.0 - 900.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0

ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	H#	WIND1 RAW CPM	WIND2 RAW CPM	WIND1 CPM	WIND1 %ERROR	WIND2 CPM	WIND2 %ERROR	LUMEX %	ELAPSED TIME
1	**-1	0.25	50.0	46572.00	52956.00	46572.00	1.85	52956.00	1.74	0.00	0.73
2	**-2	0.25	49.3	47740.00	54180.00	47740.00	1.83	54180.00	1.72	0.00	1.97
3	**-3	0.25	50.3	45272.00	51812.00	45272.00	1.88	51812.00	1.76	0.00	3.18
4	**-4	0.25	49.4	47800.00	54724.00	47800.00	1.83	54724.00	1.71	0.00	4.55
5	**-5	0.25	50.8	46324.00	53004.00	46324.00	1.86	53004.00	1.74	0.00	5.78
6	**-6	0.25	50.7	44828.00	51016.00	44828.00	1.89	51016.00	1.77	0.00	7.00
7	**-7	0.25	50.2	47660.00	54128.00	47660.00	1.83	54128.00	1.72	0.00	8.37
8	**-8	0.25	50.3	45392.00	51364.00	45392.00	1.88	51364.00	1.76	0.00	9.60
9	**-9	0.25	50.2	47560.00	54088.00	47560.00	1.83	54088.00	1.72	0.00	10.96
10	**-10	0.25	50.0	46604.00	53176.00	46604.00	1.85	53176.00	1.73	0.00	12.20
11	**-11	0.25	50.8	47200.00	53668.00	47200.00	1.84	53668.00	1.73	0.00	13.56
12	**-12	15.00	50.1	8.73	31.13	8.73	17.47	31.13	9.25	0.84	29.68

 8/4/09

ID:RN222 DONT USE 28 JUL 2009 09:27

USER:12 COMMENT:RED

PRESET TIME : 15.00

DATA CALC :	CPM	H#	:YES	SAMPLE REPEATS:	1	PRINTER	:EDIT
COUNT BLANK :	NO	IC#	: NO	REPLICATES :	1	RS232	:EDIT
TWO PHASE :	NO	AQC	: NO	CYCLE REPEATS :	1	DISK	: OFF
SCINTILLATOR:	LIQUID	LUMEX:	NO	LOW SAMPLE REJ:	0		
LOW LEVEL :	YES	HALF LIFE CORRECTION DATE:				none	

CHAN: 600.0 - 875.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0

CHAN: 0.0 - 900.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0

ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	H#	WIND1 RAW CPM	WIND2 RAW CPM	WIND1 CPM	WIND1 %ERROR	WIND2 CPM	WIND2 %ERROR	LUMEX %	ELAPSED TIME
1	57-1	10.55	51.5	952.89	1099.34	952.89	1.99	1099.34	1.86	0.03	11.15
2	57-2	10.65	49.9	942.35	1095.12	942.35	2.00	1095.12	1.85	0.02	22.86
3	57-3	10.85	50.8	924.06	1074.93	924.06	2.00	1074.93	1.85	0.02	34.78
4	57-4	10.80	49.6	926.48	1076.39	926.48	2.00	1076.39	1.85	0.02	46.78
5	57-5	10.60	50.3	944.06	1096.42	944.06	2.00	1096.42	1.86	0.02	58.44
6	57-6	10.45	49.2	958.47	1105.74	958.47	2.00	1105.74	1.86	0.02	69.98
7	57-7	10.80	51.0	928.15	1077.22	928.15	2.00	1077.22	1.85	0.02	81.97
8	57-8	10.45	50.9	959.33	1111.96	959.33	2.00	1111.96	1.86	0.02	93.50
9	57-9	10.70	51.4	937.29	1090.65	937.29	2.00	1090.65	1.85	0.03	105.29
10	57-10	10.85	52.8	924.42	1081.84	924.42	2.00	1081.84	1.85	0.02	117.34

Filename: 10:43 AMab

Instrument Type	LS 6000
Data Capture Date	28 Jul 2009 08:40:27
User Filename	C:\LSCCAPTURE\BROWN\USER12\UN072801.BSF

User Number	12
User Id	RN222
User Comments	RED

Scintillator Choice: LIQUID

Sam	Rack	Time	H#	Raw CPM1	Raw CPM2	CPM Iso1	%Err1	CPM Iso2	%Err2	LumEx
1	**-1	0.25	50.0	46572.00	52956.00	46572.00	1.85	52956.00	1.74	0.00
2	**-2	0.25	49.3	47740.00	54180.00	47740.00	1.83	54180.00	1.72	0.00
3	**-3	0.25	50.3	45272.00	51812.00	45272.00	1.88	51812.00	1.76	0.00
4	**-4	0.25	49.4	47800.00	54724.00	47800.00	1.83	54724.00	1.71	0.00
5	**-5	0.25	50.8	46324.00	53004.00	46324.00	1.86	53004.00	1.74	0.00
6	**-6	0.25	50.7	44828.00	51016.00	44828.00	1.89	51016.00	1.77	0.00
7	**-7	0.25	50.2	47660.00	54128.00	47660.00	1.83	54128.00	1.72	0.00
8	**-8	0.25	50.3	45392.00	51364.00	45392.00	1.88	51364.00	1.76	0.00
9	**-9	0.25	50.2	47560.00	54088.00	47560.00	1.83	54088.00	1.72	0.00
10	**-10	0.25	50.0	46604.00	53176.00	46604.00	1.85	53176.00	1.73	0.00
11	**-11	0.25	50.8	47200.00	53668.00	47200.00	1.84	53668.00	1.73	0.00
12	**-12	15.00	50.1	8.73	31.13	8.73	17.47	31.13	9.25	0.84

Sample Count Start Time:	28 Jul 2009 08:40:56		
Data Capture Date	28 Jul 2009 08:41:35		
User Filename	S120728---1A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	1	**-1	0.25
H#, Total Counts:	50.0	13302	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 08:42:10		
Data Capture Date	28 Jul 2009 08:42:58		
User Filename	S120728---2A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	2	**-2	0.25
H#, Total Counts:	49.3	13619	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 08:43:23		
Data Capture Date	28 Jul 2009 08:44:11		
User Filename	S120728---3A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	3	**-3	0.25
H#, Total Counts:	50.3	13010	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 08:44:45		
Data Capture Date	28 Jul 2009 08:45:25		
User Filename	S120728---4A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	4	**-4	0.25
H#, Total Counts:	49.4	13752	
Start, End, X-Axis:	0	990	Channel Number

10/4/09

Sample Count Start Time:	28 Jul 2009 08:45:59		
Data Capture Date	28 Jul 2009 08:46:42		
User Filename	S120728---5A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	5	**-5	0.25
H#, Total Counts:	50.8	13312	
Start, End, X-Axis:	0	990	Channel Number

8/4/09

Sample Count Start Time:	28 Jul 2009 08:47:12		
Data Capture Date	28 Jul 2009 08:47:56		
User Filename	S120728---6A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	6	**-6	0.25
H#, Total Counts:	50.7	12813	
Start, End, X-Axis:	0	990	Channel Number

Sample Count Start Time:	28 Jul 2009 08:48:34		
Data Capture Date	28 Jul 2009 08:49:13		
User Filename	S120728---7A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	7	**-7	0.25
H#, Total Counts:	50.2	13592	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 08:49:48		
Data Capture Date	28 Jul 2009 08:50:27		
User Filename	S120728---8A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	8	**-8	0.25
H#, Total Counts:	50.3	12904	
Start, End, X-Axis:	0	990	Channel Number

8/4/09

Sample Count Start Time:	28 Jul 2009 08:51:10		
Data Capture Date	28 Jul 2009 08:51:49		
User Filename	S120728---9A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	9	**-9	0.25
H#, Total Counts:	50.2	13586	
Start, End, X-Axis:	0	990	Channel Number

Sample Count Start Time:	28 Jul 2009 08:52:24		
Data Capture Date	28 Jul 2009 08:53:18		
User Filename	S120728---10A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	10	**-10	0.25
H#, Total Counts:	50.0	13354	
Start, End, X-Axis:	0	990	Channel Number

Sample Count Start Time:	28 Jul 2009 08:53:46		
Data Capture Date	28 Jul 2009 08:54:31		
User Filename	S120728---11A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	11	**-11	0.25
H#, Total Counts:	50.8	13477	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 08:55:08		
Data Capture Date	28 Jul 2009 09:10:30		
User Filename	S120728---12A.XLS		
	U120728---1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	12	**-12	15.00
H#, Total Counts:	50.1	474	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Filename: 10:43 AMab

Instrument Type	LS 6000
Data Capture Date	28 Jul 2009 09:11:25
User Filename	C:\LSCCAPTURE\BROWN\USER12\UN072801.BSF

User Number	12
User Id	RN222
User Comments	RED

Scintillator Choice:	LIQUID

Sam	Rack	Time	H#	Raw CPM1	Raw CPM2	CPM Iso1	%Err1	CPM Iso2	%Err2	LumEx
1	57-1	10.55	51.5	952.89	1099.34	952.89	1.99	1099.34	1.86	0.03
2	57-2	10.65	49.9	942.35	1095.12	942.35	2.00	1095.12	1.85	0.02
3	57-3	10.85	50.8	924.06	1074.93	924.06	2.00	1074.93	1.85	0.02
4	57-4	10.80	49.6	926.48	1076.39	926.48	2.00	1076.39	1.85	0.02
5	57-5	10.60	50.3	944.06	1096.42	944.06	2.00	1096.42	1.86	0.02
6	57-6	10.45	49.2	958.47	1105.74	958.47	2.00	1105.74	1.86	0.02
7	57-7	10.80	51.0	928.15	1077.22	928.15	2.00	1077.22	1.85	0.02
8	57-8	10.45	50.9	959.33	1111.96	959.33	2.00	1111.96	1.86	0.02
9	57-9	10.70	51.4	937.29	1090.65	937.29	2.00	1090.65	1.85	0.03
10	57-10	10.85	52.8	924.42	1081.84	924.42	2.00	1081.84	1.85	0.02

Sample Count Start Time:	28 Jul 2009 09:12:01		
Data Capture Date	28 Jul 2009 09:22:17		
User Filename	S12072857-1A.XLS		
	U12072857-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	1	57-1	10.55
H#, Total Counts:	51.5	11646	
Start, End, X-Axis:	0	990	Channel Number

Sample Count Start Time:	28 Jul 2009 09:23:38		
Data Capture Date	28 Jul 2009 09:34:01		
User Filename	S12072857-2A.XLS		
	U12072857-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	2	57-2	10.65
H#, Total Counts:	49.9	11708	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 09:35:21		
Data Capture Date	28 Jul 2009 09:46:26		
User Filename	S12072857-3A.XLS		
	U12072857-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	3	57-3	10.85
H#, Total Counts:	50.8	11720	
Start, End, X-Axis:	0	990	Channel Number

 8/9/09

Sample Count Start Time:	28 Jul 2009 09:47:24		
Data Capture Date	28 Jul 2009 09:57:56		
User Filename	S12072857-4A.XLS		
	U12072857-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	4	57-4	10.80
H#, Total Counts:	49.6	11690	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 09:59:15		
Data Capture Date	28 Jul 2009 10:09:36		
User Filename	S12072857-5A.XLS		
	U12072857-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	5	57-5	10.60
H#, Total Counts:	50.3	11687	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 10:10:57		
Data Capture Date	28 Jul 2009 10:21:07		
User Filename	S12072857-6A.XLS		
	U12072857-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	6	57-6	10.45
H#, Total Counts:	49.2	11606	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 10:22:35		
Data Capture Date	28 Jul 2009 10:33:08		
User Filename	S12072857-7A.XLS		
	U12072857-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	7	57-7	10.80
H#, Total Counts:	51.0	11700	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 10:34:28		
Data Capture Date	28 Jul 2009 10:44:39		
User Filename	S12072857-8A.XLS		
	U12072857-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	8	57-8	10.45
H#, Total Counts:	50.9	11680	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 10:46:00		
Data Capture Date	28 Jul 2009 10:56:26		
User Filename	S12072857-9A.XLS		
	U12072857-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	9	57-9	10.70
H#, Total Counts:	51.4	11732	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	28 Jul 2009 10:57:54		
Data Capture Date	28 Jul 2009 11:08:32		
User Filename	S12072857-10A.XLS		
	U12072857-1A.XLS		
Spectrum Type	Log Counts		
User Number	12		
User Id	RN222 DONT USE		
User Comment	RED		
Isotope Name	14C		
Scintillator	LIQUID		
Sample, Rack-Pos, Time:	10	57-10	10.85
H#, Total Counts:	52.8	11797	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Amersham International plc
Amersham Laboratories

0024

NAMAS
CALIBRATION
No. 0146

[redacted]

Description	Principal radionuclide:	Radium-226	Product code: RAY.44
			Solution number: R4/131/58
Measurement	Reference time:		1200 GMT on 10 September 1997
	Radioactive concentration of radium-226:		43.85 kilobecquerels per gram of solution
	which is equivalent to:		1.185 microcuries per gram of solution
	Mass of solution:		5.0357 grams
	Total activity of radium-226:		220.8 kilobecquerels
	which is equivalent to:		5.968 microcuries
	Recommended half life:		1600.00 years

Method of measurement:
The activity of the solution was measured in a high pressure re-entrant ionisation chamber calibrated with a large number of absolutely standardised solutions.

Accuracy

Overall uncertainty in the radioactive concentration quoted above: ± 3.7 %

Random uncertainty: ± 0.7 % Systematic uncertainty: ± 3.0 %

The limits of overall uncertainty were taken as the arithmetic sum of the uncertainty due to random variations, calculated at the 99.7% confidence level, and the estimated systematic uncertainties in the measurement.

Radionuclidic Purity

The estimated activities of any radioactive impurities found by high-resolution gamma ray spectrometry, or in any other examination of the solution, are listed below expressed as percentages of the activity of the principal radionuclide at the reference time.

Chemical Composition.

Carrier free in 0.5M HCl

Remarks

At the reference date radium-226 was shown to be in radioactive equilibrium with its daughter nuclides down the decay chain to polonium-214 and thallium-210, the precursors of lead-210. The ionisation chamber was calibrated using a standard supplied by the National Institute of Standards and Technology, Washington DC, USA.

This product meets the quality assurance requirements of the NRC Regulatory Guide 4.15 for achieving implicit NIST traceability as defined in NCRP 58 (1985).

Approved Signatory

RC-5-021-097

 8/4/09

Standard Traceability Log Rad

Source Material Info	
Parent Code:	0024
Prepared By:	Richard Kinney
Carrier Conc:	0.5 M HCL
Reference Date:	09/10/1997
Ampoule Mass (g):	5.0357 g
Uncertainty:	+/- 3.7 %
LogBook No:	RC S 021 097

A Solution Material Info	
Isotope:	Radium-226
Prepared By:	Angela Johnson
Prep Date:	01/22/1999
Verification Date:	01/23/2008
Expiration Date:	04/09/2010
Primary Code:	0024-A
Dilution(mL):	100 mL
Mass of Parent(g):	4.9958 g
Density(g/mL):	1.0044
Balance ID:	38080204

Calculations Converting parent activity to dpm/mL|dpm/g

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / (Dilution Vol) = Parent Activity (dpm/mL)

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / Density (g/mL)/ (Dilution Vol) = Parent Activity (dpm/g)

(4.9958 g) * (43.85 kBq/g) * (60000 dpm/kBq) / (100 mL) = 131439.4980 dpm/mL

(4.9958 g) * (43.85 kBq/g) * (60000 dpm/kBq) / (1.0044 g/mL) / (100 mL) = 130865.9326 dpm/g

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date
03/02/2000	Angela Johnson	10.5724	100	0024-D	13834.1 dpm/mL	02/10/2005	02/10/2006
08/23/2002	Angela Johnson	8.0371	100	0024-G	10516.626 dpm/mL	02/10/2004	02/10/2005

GEL Laboratories LLC
Version 1.0 9/18/2000

7/28/09

7/28/09

Verification for Ra-226 Standard 0024-A 09

	Isotope	Value	Uncertainty
M. Aders			
4/13/2009	0024-A	5930.765	106.9965
	0024-A	5563.022	98.7976
	0024-A	5699.674	104.7832

Mean Value (Counting) =	5731.153	97.29	**Pass**
Stdev =	185.8813485		**Rule 3 (Pass/Fail)**

Target =	5891.04
Lower Limit =	5359.390736
Upper Limit =	6102.91613
Rule 1 Pass/Fail	**Pass**
Two sigma =	371.762697
10 % of Mean =	573.1153433
Rule 2 (Pass/Fail)	**Pass**

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements
Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.
Rule 3 = The determined mean value shall be within 5% of the certificate value.

The analyst prepared three standard verification sources for Ra-226 source 0024-A by transferring portions of the degassed standard into tared glass liquid scintillation vials. Ten mL of DI Water and 10 mL mineral oil was added to each vial and the vials were shaken to mix. A Blank vial was prepared in a similar fashion using 10 mL of DI water and 10 mL of mineral oil. The standard verification vials and Background source were dark adapted for two hours and counted on LSC Green using source standard verification. Each verification source calculation was performed as follows:

Source dpm/g = (A - B)/(C)(D)
where:
A = Ver. source cpm,
B = BKG cpm,
C = System efficiency, (cpm/dpm), and
D = mass used for standard verification.

Reference RAD SOP M-001

4/14/09

Radon-222 Liquid

Filename : RN222.XLS
File type : Excel
Version # : 1.2.3

Batch : 856389
Analyst : AF1
Prep Date : 4/9/2009

Rn-222 Abundance : 1
Rn-222 Method Uncertainty : 0.1111
Geometry : 10ML MINERAL OIL/10ML SAMPLE

Spike S/N : N/A
Spike Exp Date : N/A
Spike Activity (dpm/ml): N/A
Spike Volume Added: N/A
Spike Date/Time: 4/9/2009 14:45

Pipet, 0.1 ml Stdev : +/- 0.000701 ml
Pipet, 0.5 ml Stdev : +/- 0.002564 ml

LCS S/N : 0024-A
LCS Exp Date : 7/23/2009
LCS Activity (dpm/ml): 130781.16
LCS Volume Added: 0.10

Procedure Code : LSC22RNL
Parmname : Radon-222
Required MDA : 200 pCi/L
Halflife of Radon-222 : 3.823 days

Sample Characteristics				Count raw Data				Background			
Sample ID	Sample Aliquot L	Sample Aliquot StDev. L	Sample Date/Time	Rack Position #	Counting Time (min.)	Quench#	Gross cpm	cpm	Count Time (min.)	Count Start Date/Time	Sample Decay
1201812586.1	1.0000	2.0399E-05	4/9/2009 14:45	29-2	0.5	51.1	23650	21.60	15	4/13/2009 16:30	0.478
1201812587.1	1.0000	2.0399E-05	4/9/2009 14:45	29-3	0.55	48.9	22189.09	21.60	15	4/13/2009 16:32	0.478
1201812588.1	1.0000	2.0399E-05	4/9/2009 14:45	29-4	0.5	50.2	22738	21.60	15	4/13/2009 16:34	0.478
1201812589.1	1.0000	2.0399E-05	4/9/2009 14:45	29-5	4.3	50.2	2636.28	21.60	15	4/13/2009 16:35	0.477
1201812590.1	1.0000	2.0399E-05	4/9/2009 14:45	29-6	4.45	49.7	2562.25	21.60	15	4/13/2009 16:41	0.477
1201812591.1	1.0000	2.0399E-05	4/9/2009 14:45	29-7	4.55	49.3	2521.76	21.60	15	4/13/2009 16:46	0.477
1201812592.1	1.0000	2.0399E-05	4/9/2009 14:45	29-8	15	49.6	477.4	21.60	15	4/13/2009 16:52	0.476
1201812593.1	1.0000	2.0399E-05	4/9/2009 14:45	29-9	15	50.3	486.73	21.60	15	4/13/2009 17:09	0.475
1201812594.1	1.0000	2.0399E-05	4/9/2009 14:45	29-10	15	48.5	501.67	21.60	15	4/13/2009 17:25	0.474

Calibration Data					Backgrounds		Correction Factors			
Counted on	Calibration Date	Calibration Due Date	Detector Efficiency (cpm/dpm)	Detector Efficiency Error (cpm/dpm)	Rack Position #	Count Start Date/Time	Spike Date/Time	Rn-222 Ingrowth	Rn-222 Count Correction	Net Sample Activity for MS pCi/L
LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	29-1	4/13/2009 16:14	4/9/2009 14:45	0.522	0.522	
LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	29-1	4/13/2009 16:14	4/9/2009 14:45	0.522	0.522	
LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	29-1	4/13/2009 16:14	4/9/2009 14:45	0.522	0.522	
LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	29-1	4/13/2009 16:14	4/9/2009 14:45	0.523	0.523	
LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	29-1	4/13/2009 16:14	4/9/2009 14:45	0.523	0.523	
LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	29-1	4/13/2009 16:14	4/9/2009 14:45	0.523	0.523	
LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	29-1	4/13/2009 16:14	4/9/2009 14:45	0.524	0.524	
LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	29-1	4/13/2009 16:14	4/9/2009 14:45	0.525	0.525	
LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	29-1	4/13/2009 16:14	4/9/2009 14:45	0.526	0.526	

4/14/09

Notes:
1 - Results are decay corrected to Sample Date/Time
2 - Reference date for Spike Activity (dpm/ml) is the batch Prep Date
3 - Spike Nominals are decay corrected to Sample Date/Time

Results Decision Level pCi/L	Critical Level pCi/L	MDA pCi/L	Sample Act. Conc. pCi/L	Sample Act. Error pCi/L	Net Count Rate CPM	Net Count Rate Error CPM	2 SIGMA Counting Uncertainty pCi/L	2 SIGMA Total Prop. Uncertainty pCi/L	Sample QC	Sample Type	RPD	RER	Nominal pCi/L	Recovery
3.9075	2.7587	7.0234	**5930.7645**	0.0121	23628.4000	217.4889	106.9965	1299.1505		LCS			5891.0434	100.7%
3.7309	2.6341	6.6370	**5563.0219**	0.0120	22167.4900	200.8612	98.7976	1218.4676		LCS			5891.0434	94.4%
3.9060	2.7576	7.0207	**5699.6739**	0.0123	22716.4000	213.2544	104.8732	1248.6996		LCS			5891.0434	96.8%
1.4856	1.0489	2.2727	**655.7702**	0.0124	2614.6800	24.7897	12.1860	143.6780		LCS			612.8784	107.0%
1.4651	1.0344	2.2377	**636.7951**	0.0123	2540.6500	24.0255	11.8028	139.5180		LCS			612.8784	103.9%
1.4517	1.0249	2.2150	**626.2343**	0.0123	2500.1600	23.5727	11.5727	137.2013		LCS			612.8784	102.2%
0.9892	0.6984	1.4467	**114.0215**	0.0149	455.8000	5.7677	2.8280	25.0520		LCS			108.5740	105.0%
0.9873	0.6970	1.4440	**116.1380**	0.0148	465.1300	5.8214	2.8489	25.5135		LCS			108.5740	107.0%
0.9855	0.6957	1.4413	**119.6448**	0.0146	480.0700	5.9063	2.8851	26.2784		LCS			108.5740	110.2%

ECL 6/14/09

SAM NO	PUS	TIME MIN	HH	WIND1 CPM	WIND1 %ERROR	WIND2 CPM	WIND2 %ERROR	LUMEX %	ELAPSED TIME
[illegible]	[illegible]	[illegible]	[illegible]	10.33	[illegible]	21.60	[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]	[illegible]	21172.00	[illegible]	23650.00	[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]	[illegible]	19761.32	[illegible]	22189.09	[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]	[illegible]	20378.00	[illegible]	22738.00	[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]	[illegible]	2350.47		2635.28	[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]	[illegible]	2261.57	[illegible]	2562.25	[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]	[illegible]	2220.22	[illegible]	2521.76	[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]	[illegible]	410.33	[illegible]	477.40	[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]	[illegible]	424.07	[illegible]	436.73	[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]	[illegible]	438.53	[illegible]	501.67	[illegible]	[illegible]	[illegible]

Q025-A Q024-A

Q025-E

8/4/09

KLCA 4/14/09

Sample Count Start Time:	13 Apr 2009 16:14:45		
Data Capture Date:	13 Apr 2009 16:30:02		
User Filename:	S16041329-1A.WK1		
	U16041329-1A.WK1		
Spectrum Type	Log Counts		
User Number:	16		
User Id:	RN-222		
User Comment:	GREEN		
Isotope Name:	14C		
Scintillator:	LIQUID		
Sample, Rack-Pos, Time:	1	29-1	15.00
H#, Total Counts:	48.1	450	
Start, End, X-Axis:	0	990	Channel Number

Sample Count Start Time:	13 Apr 2009 16:30:50		
Data Capture Date:	13 Apr 2009 16:31:39		
User Filename:	S16041329-2A.WK1		
	U16041329-1A.WK1		
Spectrum Type	Log Counts		
User Number:	16		
User Id:	RN-222		
User Comment:	GREEN		
Isotope Name:	14C		
Scintillator:	LIQUID		
Sample, Rack-Pos, Time:	2	29-2	0.50
H#, Total Counts:	51.1	12433	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

Sample Count Start Time:	13 Apr 2009 16:32:27		
Data Capture Date:	13 Apr 2009 16:33:19		
User Filename:	S16041329-3A.WK1		
	U16041329-1A.WK1		
Spectrum Type	Log Counts		
User Number:	16		
User Id:	RN-222		
User Comment:	GREEN		
Isotope Name:	14C		
Scintillator:	LIQUID		
Sample, Rack-Pos, Time:	3	29-3	0.55
H#, Total Counts:	48.9	12867	
Start, End, X-Axis:	0	990	Channel Number

 LL 8/4/09

Sample Count Start Time:	13 Apr 2009 16:34:09		
Data Capture Date:	13 Apr 2009 16:34:57		
User Filename:	S16041329-4A.WK1		
	U16041329-1A.WK1		
Spectrum Type	Log Counts		
User Number:	16		
User Id:	RN-222		
User Comment:	GREEN		
Isotope Name:	14C		
Scintillator:	LIQUID		
Sample, Rack-Pos, Time:	4	29-4	0.50
H#, Total Counts:	50.2	11972	
Start, End, X-Axis:	0	990	Channel Number

 8/4/09

General Engineering Laboratories
Calibration Source Preparation Sheet

Applicable SOP Number: GL-RAD-A-007

Isotope: Rn-222

Date Standards Prepared: 2/22/06

Cocktail Type Used: Mineral Oil Scintillato

Standard ID: 0024-A

Matrix of Vial/Planchett: 10 mL DI water
10 mL Mineral Oil

Amount Used (g or mL): 0.1

Standard Activity (DPM/g or mL): 131439.4980

Type of Scintillation Vial: glass

Reference Date: 9/10/97

Pipette ID Used: 1429303

Expiration Date: 12/15/06

Balance ID Used: N/A

Residue/Carrier Agent: N/A

Quenching Agent: N/A

	Standard Number	Quenching Vol (uL)/ Residue Volume(mL)	Initial Wt. (g)	Final Wt. (g)	Net Wt. (mg)
	c1				
	c2				
	c3				
	c4				
	c5				
	c6				
	c7				
	c8				
	c9				
	c10				
	c11				

Prepared By: Amanda L. Fehr Date: 2/22/06

Reviewed By: [illegible] Date: 3/30/06

Rev 1 RLM 9/10/97

 8/4/09

Rec'd 8-21-00

Nycomed Amersham plc
Amersham Laboratories

0299

NAMAS
CALIBRATION
No. 0146

[redacted]

ISSUED	Nycomed Amersham plc Radiation & Radioactivity Calibration Laboratory Amersham Laboratories White Lion Road Amersham Buckinghamshire HP7 9LL	ISSUED FOR:	AEA Technology plc Isotrak Amersham Laboratories White Lion Road Amersham Buckinghamshire HP7 9LL

Description Principal radionuclide: Radium-226

Product code: RAY44
Solution number: R4/131/89

Measurement Reference time: 1200 GMT on 15 December 1999

Nuclear data Nuclear data quoted on this certificate are taken from the Joint European File, Version 2.2.

Expression of uncertainties The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k = 2.00, which for a *t*-distribution with $\nu_{eff} = \infty$ effective degrees of freedom corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Unless indicated, all other uncertainties are expressed at the confidence level associated with one standard uncertainty.

The format used for the uncertainties in the values of radionuclidic purity is illustrated in the following examples;

6.5(21)	=	6.5 ± 2.1
6.54(21)	=	6.54 ± 0.21
6.543(21)	=	6.543 ± 0.021

Approved Signatory

Date of issue 17th December 1999

Nycomed Amersham

Standard Traceability Log Rad

Source Material Info	
Parent Code:	0299
Prepared By:	Angela Johnson
Carrier Conc:	0.5 M HCL
Reference Date:	12/15/1999
Ampoule Mass (g):	5.0368 g
Uncertainty:	+/- 2.5 %
LogBook No:	RC S 027 128

A Solution Material Info	
Isotope:	Radium-226
Prepared By:	Angela Johnson
Prep Date:	09/15/2000
Verification Date:	01/23/2008
Expiration Date:	01/23/2009
Primary Code:	0299-A
Dilution(mL):	100 mL
Mass of Parent(g):	4.6634 g
Density(g/mL):	1.0012
Balance ID:	

Calculations Converting parent activity to dpm/mL|dpm/g

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / (Dilution Vol) = Parent Activity (dpm/mL)

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / Density (g/mL)/ (Dilution Vol) = Parent Activity (dpm/g)

(4.6634 g) * (43.75 kBq/g) * (60000 dpm/kBq) / (100 mL) = 122414.2500 dpm/mL

(4.6634 g) * (43.75 kBq/g) * (60000 dpm/kBq) / (1.0012 g/mL) / (100 mL) = 122273.3377 dpm/g

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date
08/26/2003	Angela Johnson	1.9909	100	0299-E	2434.34 dpm/mL	11/04/2004	11/04/2005
08/26/2003	Angela Johnson	1.9872	100	0299-F	2429.82 dpm/mL	08/26/2004	08/26/2005
04/05/2005	Amanda Fehr	5.0018	250	0299-G	2446.3471 dpm/mL	04/02/2008	04/02/2009

GEL Laboratories LLC
Version 1.0 9/18/2000

 8/4/09

Verification for Ra-226 Standard 0299-G

4/2/2008 D. Roy	Isotope	Detector CPM	BKG CPM	NET CPM	Detector Eff	Standard Mass. Used (G)	Source DPM/G
	0299-G N1	2536.9600	52.4000	2484.5600	1.917186	0.5057	2562.667649
	0299-G N2	2520.2500	52.4000	2467.8500	1.917186	0.5056	2545.935781
	0299-G N3	2532.5000	52.4000	2480.1000	1.917186	0.5042	2565.677715
Mean Value (Counting) =	2558.093715		**104.944421**	**Pass**		Average =	2558.093715
Stdev =	10.63610098		0.00415782	**Rule 3 (Pass/Fail)**			

Certificate Value =	2437.6	dpm/mL
Lower Limit =	2536.821513	dpm/mL
Upper Limit =	2579.365917	dpm/mL
Rule 1 Pass/Fail	**Fail**	***exception taken due to full recovery of standard**
Two sigma =	21.27220197	dpm/mL
10 % of Mean =	255.8093715	dpm/mL
Rule 2 (Pass/Fail)	**Pass**	

Verification Rules

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements

Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.

Rule 3 = The determined mean value shall be within 10% of the certificate value.

The analyst prepared three standard verification sources for Ra-226 source 0299-G by transferring portions of the standard into tared glass liquid scintillation vials. One mL of DI Water and ten mLs of Ready Gel liquid scintillation cocktail was added to each vial and the vials were shaken to mix. A Blank vial was prepared in a similar fashion using 1 mL of DI water and 10 mL of Ready Gel cocktail. The standard verification vials and Background source were dark adapted for two hours and counted on LSC Gold for Radium source standard verification. The Ra-226 efficiency calibration which was used for verification calculations was performed on 4/02/08 using source 0024-A (Ra-226). Calibration data is recorded in this logbook under Ra-226 0024. Each verification source calculation was performed as follows:

Source dpm/g = (A - B)/(C)(D)

where:

A = Ver. source cpm,
B = BKG cpm,
C = System efficiency, (cpm/dpm), and
D = mass used for standard verification.

RAD SOP M-001

4/9/08

Mary Jo. Adams 4/9/08

David Roy 4/10/08

ID: TOTAL ACTIVITY

2 APR 2008 22:17

USER:11 COMMENT:GOLD

PRESET TIME :	5.00						
DATA CALC :	CPM	H#	:YES	SAMPLE REPEATS:	1	PRINTER	: STD
COUNT BLANK :	NO	IC#	: NO	REPLICATES :	1	RS232	:EDIT
TWO PHASE :	NO	AQC	: NO	CYCLE REPEATS :	1	DISK	: OFF
SCINTILLATOR:	LIQUID	LUMEX	:YES	LOW SAMPLE REJ:	0		
LOW LEVEL :	NO	HALF LIFE CORRECTION DATE:				none	

CHAN: 0.0 - 990.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0
CHAN: 0.0 - 1000.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0

	SAM NO	POS	TIME MIN	H#	WIND1 CPM	WIND1 %ERROR	WIND2 CPM	WIND2 %ERROR	LUMEX %	ELAPSED TIME
0024-A	1	13-1	5.00	111.9	48.60	12.86	51.60	12.48	0.89	5.59
	2	13-2	0.40	114.1	25475.00	1.98	25475.00	1.98	0.00	6.94
	3	13-3	0.40	114.3	25432.50	1.98	25432.50	1.98	0.00	8.34
	4	13-4	0.45	114.1	24786.67	1.89	24786.67	1.89	0.00	9.80
0299-G	5	13-5	5.00	112.3	52.40	12.36	54.80	12.08	0.57	15.89
	6	13-6	3.95	121.6	2536.96	2.00	2538.23	2.00	0.07	20.89
	7	13-7	4.00	120.4	2520.25	1.99	2523.00	1.99	0.05	25.97
	8	13-8	4.00	120.0	2532.50	1.99	2534.25	1.99	0.05	31.04

Sample Count Start Time:	2 Apr 2008 22:28:46		
Data Capture Date:	4/2/2008 22:34:01		
User Filename:	S11040213-5A.WK1		
	U11040213-1A.WK1		
Spectrum Type	Log Counts		
User Number:	11		
User Id:	TOTAL ACTIVITY		
User Comment:	GOLD		
Isotope Name:	14C		
Scintillator:	LIQUID		
Sample, Rack-Pos, Time:	5	13-5	5.00
H#, Total Counts:	112.3	66	
Start, End, X-Axis:	0	990	Channel Number

DM 4/10/08

LH 8/4/09

Sample Count Start Time:	2 Apr 2008 22:34:49		
Data Capture Date:	4/2/2008 22:39:04		
User Filename:	S11040213-6A.WK1		
	U11040213-1A.WK1		
Spectrum Type	Log Counts		
User Number:	11		
User Id:	TOTAL ACTIVITY		
User Comment:	GOLD		
Isotope Name:	14C		
Scintillator:	LIQUID		
Sample, Rack-Pos, Time:	6	13-6	3.95
H#, Total Counts:	121.6	7906	
Start, End, X-Axis:	0	990	Channel Number

Sample Count Start Time:	2 Apr 2008 22:39:51		
Data Capture Date:	4/2/2008 22:44:07		
User Filename:	S11040213-7A.WK1		
	U11040213-1A.WK1		
Spectrum Type	Log Counts		
User Number:	11		
User Id:	TOTAL ACTIVITY		
User Comment:	GOLD		
Isotope Name:	14C		
Scintillator:	LIQUID		
Sample, Rack-Pos, Time:	7	13-7	4.00
H#, Total Counts:	120.4	7901	
Start, End, X-Axis:	0	990	Channel Number

4/10/08

8/4/09

Sample Count Start Time:	2 Apr 2008 22:44:55		
Data Capture Date:	4/2/2008 22:49:12		
User Filename:	S11040213-8A.WK1		
	U11040213-1A.WK1		
Spectrum Type	Log Counts		
User Number:	11		
User Id:	TOTAL ACTIVITY		
User Comment:	GOLD		
Isotope Name:	14C		
Scintillator:	LIQUID		
Sample, Rack-Pos, Time:	8	13-8	4.00
H#, Total Counts:	120.0	7908	
Start, End, X-Axis:	0	990	Channel Number

General Engineering Laboratories
Verification Source Preparation Sheet

Applicable SOP Number: GL-RAD-A-007

Isotope: Rn222

Date Standards Prepared: 2/22/06

Cocktail Type Used: Mineral Oil Scintilator

Standard ID: 0299-G

Matrix of Vial/Planchett: 10 mL DI water
10 mL Mineral Oil

Amount Used (g or mL): ~~0.5~~ 0.1

Standard Activity (DPM/g or mL): 2446.3471

Type of Scintillation Vial: glass

Reference Date: 12/15/99

Pipette ID Used: 3158763

Expiration Date: 12/15/06

Balance ID Used: N/A

Residue/Carrier Agent: N/A

Quenching Agent: N/A

	Standard Number	Quenching Vol (uL)/ Residue Volume(mL)	Initial Wt. (g)	Final Wt. (g)	Net Wt. (mg)
	1				
	2				
	3				
	4				
	5				
	6				
	7				
	8				
	9				
	10				
	11				

Prepared By: Amanda L. Fehr Date: 2/22/06

Reviewed By: [illegible] Date: 3/30/06

Rev 1 RLM 9/10/97

[illegible] 3/4/09

Continuing Calibration Data

Liquid Scintillation Counter Checks for 9-JUL-2010

Short Name	Parmname	Run Time	Count Time	Counts	CPM	Stdev	Status	Comments
LSCBLUE	Carbon-14	02:45	.95		42149	0.54		
	Carbon-14	02:47	1		42015	-0.19		
	Tritium	02:48	.85		61648	-1.5		
	Tritium	02:49	.85		61871	-1.1		
	BKG	02:51	30		21	-1.2		
	BKG	03:22	30		21	-0.54		
LSCBROWN	Carbon-14	04:10	1		42173	-0.5		
	Carbon-14	04:11	.95		42310	0.14		
	Tritium	04:13	.85		61712	-1.6		
	Tritium	04:15	.85		61990	-0.61		
	BKG	04:17	30		23	0.78		
	BKG	04:48	30		23	-0.23		
LSCGOLD	Carbon-14	02:51	.95		42983	-1.5		
	Carbon-14	02:52	.95		43101	-1		
	Tritium	02:54	.75		83128	-2		
	Tritium	02:55	.75		82624	-2.9		
	BKG	02:57	30		21	1.1		
	BKG	03:27	30		18	-2.2		
LSCGREEN	Carbon-14	02:51	.95		42131	-0.05		
	Carbon-14	02:52	.95		42223	0.41		
	Tritium	02:54	.9		58282	-2.9		
	Tritium	02:55	.9		58851	-1.2		
	BKG	02:57	30		20	-0.21		
	BKG	03:28	30		23	2.86		
LSCORANGE	Carbon-14	13:42	1.01245		98701	1.17		
	Tritium	13:46	1.012333		132129	0.31		
	BKG	14:18	30.01243		11	-0.6		
	Carbon-14	14:20	1.012433		98363	0.21		
	Tritium	14:21	1.012433		131919	-0.2		
	BKG	14:52	30.01243		12	1.09		
LSCPINK	Carbon-14	01:37	1.012433		93542	-0.21		
	Tritium	01:40	1.012433		55568	-0.04		

	BKG	02:13	30.01243	7.25	-0.34	
	Carbon-14	02:14	1.012433	93476	-0.38	
	Tritium	02:16	1.012433	55022	-0.62	
	BKG	02:46	30.01243	7.55	0.11	
	Carbon-14	11:43	1.012433	92966	-1.7	
	Tritium	11:46	1.012333	55447	-0.17	
	BKG	12:19	30.01243	7.12	-0.54	
	Carbon-14	12:20	1.012433	93437	-0.48	
	Tritium	12:22	1.012433	54601	-1.1	
	BKG	12:52	30.01243	8.13	1	
	Carbon-14	13:16	1.012433	93629	0.02	
	Tritium	13:19	1.012433	56048	0.46	
	BKG	13:52	30.01243	7.59	0.18	
	Carbon-14	13:53	1.012433	93500	-0.32	
	Tritium	13:55	1.012433	56059	0.47	
	BKG	14:25	30.01243	8.84	2.08	
	Carbon-14	16:08	1.01245	93997	1	
	Tritium	16:12	1.012333	55816	0.22	
	BKG	16:44	30.01243	7.93	0.69	
	Carbon-14	16:46	1.012433	93337	-0.75	
	Tritium	16:47	1.012433	55518	-0.1	
	BKG	17:30	30.01442	6.88	-0.91	
LSCPURPLE	Carbon-14	03:12	.55	93103	-4	~~Outside Three Sigma~~ JP 7/13/10
	Carbon-14	03:12	.55	94418	-0.71	cleared by rerun
	Tritium	03:13	.95	65304	-1.7	
	Tritium	03:15	.95	65240	-1.9	
	BKG	03:17	30	28	-0.93	
	BKG	03:47	30	28	-0.82	
LSCRED	Carbon-14	03:01	.95	42144	-0.75	
	Carbon-14	03:02	.95	42307	-0.02	
	Tritium	03:04	.8	64856	-1.7	
	Tritium	03:05	.8	64228	-3.5	~~Outside Three Sigma~~ JP 7/13/10
	BKG	03:07	30	25	3	cleared by rerun
	BKG	03:37	30	23	1.22	
	Carbon-14	10:21	.95	42380	0.32	
	Carbon-14	10:22	.95	42632	1.46	
	Tritium	10:24	.8	64863	-1.7	
	Tritium	10:25	.8	64677	-2.2	

	BKG	10:27	30	24	1.87
	BKG	10:58	30	21	-0.87
LSCSILVER	Carbon-14	02:31	.45	95847	1.1
	Carbon-14	02:32	.45	94869	-1.2
	Tritium	02:33	.8	69023	-1.6
	Tritium	02:34	.8	69056	-1.5
	BKG	02:35	30	26	1.2
	BKG	03:06	30	24	-0.3
LSCYELLOW	Carbon-14	02:29	1.717	97263	-1.3
	Carbon-14	02:32	1.733	96907	-3
	Tritium	02:34	2.117	127631	-1
	Tritium	02:37	2.133	126962	-2.4
	BKG	02:40	30	13	-2
	BKG	03:11	30	13	-0.8

Reviewed by Lynnette Pace

Date 7/13/10

GEL Laboratories LLC

Liquid Scintillation Counter Checks for 11-JUL-2010

Short Name	Parmname	Run Time	Count Time	Counts	CPM	Stdev	Status	Comments
LSCBLUE	Carbon-14	09:36	.95		42360	1.69		
	Carbon-14	09:38	.95		42261	1.15		
	Tritium	09:39	.85		61427	-1.9		
	Tritium	09:40	.85		61903	-1.1		
	BKG	09:42	30		21	-1.2		
	BKG	10:13	30		22	0		
LSCBROWN	Carbon-14	05:32	.95		42468	0.88		
	Carbon-14	05:34	.95		42315	0.16		
	Tritium	05:36	.85		61916	-0.87		
	Tritium	05:37	.85		62394	0.83		
	BKG	05:39	30		22	-0.75		
	BKG	06:10	30		21	-1.8		
LSCGOLD	Carbon-14	03:29	.95		43437	0.42		
	Carbon-14	03:30	.95		43317	-0.1		
	Tritium	03:32	.75		83284	-1.7		
	Tritium	03:33	.75		82482	-3.2	~~Outside Three Sigma~~ JP 7/13/10	
	BKG	03:35	30		20	0.18	cleared by rerun	
	BKG	04:05	30		20	-0.87		
	Tritium	14:32	.75		84332	0.23		
	Tritium	14:33	.75		83770	-0.82		
	Tritium	14:45	.75		83718	-0.91		
	Tritium	14:46	.75		83206	-1.9		
LSCGREEN	Carbon-14	02:36	.95		42210	0.35		
	Carbon-14	02:37	.95		42260	0.6		
	Tritium	02:39	.9		58885	-1.1		
	Tritium	02:40	.85		59559	0.97		
	BKG	02:42	30		22	1.93		
	BKG	03:13	30		22	1.93		
LSCORANGE	Carbon-14	08:43	1.012433		98308	0.05		
	Tritium	08:47	1.012433		131140	-2.1		
	BKG	09:19	30.01243		12	0.81		
	Carbon-14	09:21	1.012433		97453	-2.4		

	Tritium	09:22	1.012333	132332	0.81
	BKG	09:53	30.01243	12	0.66
LSCRED	Carbon-14	05:38	.95	42535	1.02
	Carbon-14	05:40	.95	42377	0.3
	Tritium	05:41	.8	64981	-1.4
	Tritium	05:42	.8	65077	-1.1
	BKG	05:44	30	23	0.62
	BKG	06:15	30	24	1.8
LSCSILVER	Carbon-14	02:30	.45	95216	-0.39
	Carbon-14	02:31	.45	95809	1.02
	Tritium	02:32	.75	69249	-1
	Tritium	02:33	.75	69200	-1.1
	BKG	02:35	30	24	-0.42
	BKG	03:05	30	23	-1.7
LSCYELLOW	Carbon-14	08:18	1.717	97441	-0.54
	Carbon-14	08:20	1.717	97442	-0.54
	Tritium	08:23	2.133	126856	-2.7
	Tritium	08:26	2.133	126729	-2.9
	BKG	08:29	30	14	0.71
	BKG	09:00	30	13	-1.3

Reviewed by Lynsey Pace

Date 7/13/10

GEL Laboratories LLC

Background and Efficiency Data

● Denotes Outlier

Quality Control Charts

Data used for LSC Liquids 14-JUL-2010

Radon-222 BLANK: Limits LCL = 0 UCL = .1

Batch ID	Samp ID	Run Date	Status	Value	Deviation	Units	Mean	LCL	LWL	UWL	UCL	Stdev
967258	1202076014	19-MAR-2010 22:52	DONE	0	-0.26	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
968204	1202078478	23-MAR-2010 19:24	DONE	0	0.2	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
968303	1202078708	23-MAR-2010 20:28	DONE	0	-0.26	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
968961	1202080315	24-MAR-2010 19:46	DONE	0	0.98	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
968842	1202079948	24-MAR-2010 20:50	DONE	0	-0.27	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
969311	1202081133	25-MAR-2010 19:38	DONE	0	-2	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
969700	1202082146	26-MAR-2010 19:19	DONE	0	1.1	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
970058	1202082955	29-MAR-2010 14:36	DONE	0	-0.67	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
970475	1202083969	30-MAR-2010 23:53	DONE	0	-0.27	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
970477	1202083973	31-MAR-2010 01:29	DONE	0	-2	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
970839	1202084974	01-APR-2010 01:09	DONE	0	-0.02	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
971267	1202086034	01-APR-2010 23:26	DONE	0	-0.63	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
972847	1202089822	07-APR-2010 17:40	DONE	0	-1	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
973312	1202090932	08-APR-2010 21:18	DONE	0	1.2	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
974603	1202094120	13-APR-2010 22:30	DONE	0	-0.68	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
975286	1202095814	15-APR-2010 21:39	DONE	0	0.7	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
975561	1202096481	16-APR-2010 18:27	DONE	0	0.55	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
976372	1202098554	20-APR-2010 20:29	DONE	0	0.82	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
976828	1202099677	21-APR-2010 23:22	DONE	0	-0.63	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
977370	1202100756	22-APR-2010 22:21	DONE	0	-0.82	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
977678	1202101490	23-APR-2010 20:33	DONE	0	0.56	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
978565	1202103618	27-APR-2010 22:30	DONE	0	-0.08	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
978981	1202104623	28-APR-2010 21:57	DONE	0	0.71	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
979717	1202106349	29-APR-2010 21:22	DONE	0	1.2	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
980337	1202107710	03-MAY-2010 22:34	DONE	0	-0.02	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
981849	1202111217	07-MAY-2010 07:17	DONE	0	-1	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
982817	1202113774	11-MAY-2010 22:52	DONE	0	0.34	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
983453	1202115457	12-MAY-2010 04:05	DONE	0	0.18	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
983787	1202116236	13-MAY-2010 01:15	DONE	0	-1	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
984139	1202117009	13-MAY-2010 23:13	DONE	0	-0.27	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
984715	1202118127	17-MAY-2010 05:13	DONE	0	2.7	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
985996	1202121045	20-MAY-2010 19:37	DONE	0	1.4	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
988956	1202128071	27-MAY-2010 18:25	DONE	0	-0.51	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
989508	1202129294	01-JUN-2010 19:29	DONE	0	1.7	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
991377	1202133731	05-JUN-2010 17:53	DONE	0	-0.68	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
992256	1202135973	09-JUN-2010 06:40	DONE	0	0.46	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
993157	1202138140	10-JUN-2010 20:58	DONE	0	-0.56	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
1000953	1202157537	09-JUL-2010 16:42	DONE	0	-0.86	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02
1001621	1202159127	09-JUL-2010 22:07	DONE	0	0.53	pCi/mL	00	-0.045	-0.029	0.04	0.05	0.02

Radon-222 DUP: Limits LCL = 0 UCL = 20

Batch ID	Samp ID	Run Date	Status	Value	Deviation	Units	Mean	LCL	LWL	UWL	UCL	Stdev
968204	1202078479	23-MAR-2010 19:40	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
968303	1202078709	23-MAR-2010 20:44	DONE	2	-0.39	percent	9.81	0	-29	48.9	20.0	19.6
968961	1202080316	24-MAR-2010 20:02	DONE	47	1.9	percent	9.81	0	-29	48.9	20.0	19.6
968842	1202079949	24-MAR-2010 21:06	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6

969311	1202081134	25-MAR-2010 19:54	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
969700	1202082147	26-MAR-2010 19:35	DONE	42	1.7	percent	9.81	0	-29	48.9	20.0	19.6
970058	1202082956	29-MAR-2010 14:52	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
970475	1202083970	31-MAR-2010 00:09	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
970477	1202083974	31-MAR-2010 01:46	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
970839	1202084975	01-APR-2010 01:25	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
971267	1202086035	01-APR-2010 23:42	DONE	50	2	percent	9.81	0	-29	48.9	20.0	19.6
972847	1202089823	07-APR-2010 17:56	DONE	2	-0.38	percent	9.81	0	-29	48.9	20.0	19.6
973312	1202090933	08-APR-2010 21:34	DONE	10	00	percent	9.81	0	-29	48.9	20.0	19.6
974603	1202094121	13-APR-2010 22:46	DONE	16	0.3	percent	9.81	0	-29	48.9	20.0	19.6
975286	1202095815	15-APR-2010 21:56	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
975561	1202096482	16-APR-2010 18:43	DONE	2	-0.4	percent	9.81	0	-29	48.9	20.0	19.6
976372	1202098555	20-APR-2010 20:46	DONE	17	0.35	percent	9.81	0	-29	48.9	20.0	19.6
976828	1202099678	21-APR-2010 23:38	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
977370	1202100757	22-APR-2010 22:37	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
977678	1202101491	23-APR-2010 20:49	DONE	0	-0.48	percent	9.81	0	-29	48.9	20.0	19.6
978565	1202103619	27-APR-2010 22:46	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
978981	1202104624	28-APR-2010 22:13	DONE	97	4.5	percent	9.81	0	-29	48.9	20.0	19.6
979717	1202106350	29-APR-2010 21:38	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
980337	1202107711	03-MAY-2010 22:50	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
981849	1202111218	07-MAY-2010 07:33	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
982817	1202113775	12-MAY-2010 02:05	DONE	5	-0.27	percent	9.81	0	-29	48.9	20.0	19.6
982817	1202113776	12-MAY-2010 04:07	DONE	19	0.45	percent	9.81	0	-29	48.9	20.0	19.6
983453	1202115458	12-MAY-2010 04:21	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
983787	1202116237	13-MAY-2010 01:31	DONE	5	-0.24	percent	9.81	0	-29	48.9	20.0	19.6
984139	1202117010	13-MAY-2010 23:29	DONE	10	0.01	percent	9.81	0	-29	48.9	20.0	19.6
984715	1202118128	17-MAY-2010 05:29	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
985996	1202121046	20-MAY-2010 19:53	DONE	1	-0.47	percent	9.81	0	-29	48.9	20.0	19.6
988956	1202128072	27-MAY-2010 18:41	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
989508	1202129295	01-JUN-2010 19:45	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
991377	1202133732	05-JUN-2010 18:24	DONE	3	-0.37	percent	9.81	0	-29	48.9	20.0	19.6
992256	1202135974	09-JUN-2010 06:57	DONE	12	0.09	percent	9.81	0	-29	48.9	20.0	19.6
993157	1202138141	10-JUN-2010 21:14	DONE	0	-0.5	percent	9.81	0	-29	48.9	20.0	19.6
1000953	1202157538	09-JUL-2010 17:13	DONE	32	1.1	percent	9.81	0	-29	48.9	20.0	19.6
1001621	1202159128	09-JUL-2010 22:24	DONE	13	0.17	percent	9.81	0	-29	48.9	20.0	19.6

Radon-222 LCS: Limits LCL = 75 UCL = 125

Batch ID	Samp ID	Run Date	Status	Value	Deviation	Units	Mean	LCL	LWL	UWL	UCL	Stdev
967258	1202076016	22-MAR-2010 13:48	DONE	98	-0.24	percent	100	75.0	83.3	117	125	8.33
966794	1202074887	22-MAR-2010 14:32	DONE	112	1.4	percent	100	75.0	83.3	117	125	8.33
968303	1202078710	27-MAR-2010 09:01	DONE	78	-3	percent	100	75.0	83.3	117	125	8.33
969311	1202081136	29-MAR-2010 05:09	DONE	102	0.28	percent	100	75.0	83.3	117	125	8.33
968204	1202078481	29-MAR-2010 22:27	DONE	102	0.25	percent	100	75.0	83.3	117	125	8.33
969700	1202082149	29-MAR-2010 23:15	DONE	95	-0.6	percent	100	75.0	83.3	117	125	8.33
968842	1202079950	29-MAR-2010 23:48	DONE	99	-0.17	percent	100	75.0	83.3	117	125	8.33
968961	1202080318	30-MAR-2010 00:37	DONE	101	0.1	percent	100	75.0	83.3	117	125	8.33
970058	1202082958	01-APR-2010 18:26	DONE	103	0.38	percent	100	75.0	83.3	117	125	8.33
970477	1202083976	02-APR-2010 04:09	DONE	107	0.84	percent	100	75.0	83.3	117	125	8.33
970475	1202083971	02-APR-2010 04:56	DONE	111	1.3	percent	100	75.0	83.3	117	125	8.33
970839	1202084977	03-APR-2010 06:01	DONE	110	1.2	percent	100	75.0	83.3	117	125	8.33

971267	1202086037	05-APR-2010 05:01	DONE	105	0.61	percent	100	75.0	83.3	117	125	8.33
972847	1202089825	11-APR-2010 14:23	DONE	107	0.88	percent	100	75.0	83.3	117	125	8.33
973312	1202090935	11-APR-2010 14:24	DONE	90	-1	percent	100	75.0	83.3	117	125	8.33
974603	1202094123	16-APR-2010 16:01	DONE	107	0.86	percent	100	75.0	83.3	117	125	8.33
975561	1202096484	19-APR-2010 10:10	DONE	109	1.1	percent	100	75.0	83.3	117	125	8.33
975286	1202095817	19-APR-2010 10:58	DONE	109	1.1	percent	100	75.0	83.3	117	125	8.33
976372	1202098557	23-APR-2010 08:01	DONE	106	0.75	percent	100	75.0	83.3	117	125	8.33
976828	1202099680	24-APR-2010 07:24	DONE	98	-0.29	percent	100	75.0	83.3	117	125	8.33
977678	1202101493	26-APR-2010 18:55	DONE	107	0.85	percent	100	75.0	83.3	117	125	8.33
977370	1202100759	27-APR-2010 10:32	DONE	101	0.07	percent	100	75.0	83.3	117	125	8.33
978565	1202103621	30-APR-2010 03:16	DONE	102	0.26	percent	100	75.0	83.3	117	125	8.33
978981	1202104626	01-MAY-2010 17:11	DONE	99	-0.15	percent	100	75.0	83.3	117	125	8.33
979717	1202106352	02-MAY-2010 18:35	DONE	94	-0.7	percent	100	75.0	83.3	117	125	8.33
980337	1202107713	06-MAY-2010 03:55	DONE	110	1.2	percent	100	75.0	83.3	117	125	8.33
981849	1202111220	10-MAY-2010 06:54	DONE	107	0.78	percent	100	75.0	83.3	117	125	8.33
982817	1202113777	14-MAY-2010 10:43	DONE	115	1.8	percent	100	75.0	83.3	117	125	8.33
983453	1202115460	14-MAY-2010 19:39	DONE	118	2.1	percent	100	75.0	83.3	117	125	8.33
984139	1202117012	17-MAY-2010 08:55	DONE	115	1.8	percent	100	75.0	83.3	117	125	8.33
984715	1202118130	18-MAY-2010 00:59	DONE	102	0.21	percent	100	75.0	83.3	117	125	8.33
983787	1202116239	19-MAY-2010 06:13	DONE	99	-0.08	percent	100	75.0	83.3	117	125	8.33
985996	1202121048	23-MAY-2010 21:41	DONE	113	1.5	percent	100	75.0	83.3	117	125	8.33
988956	1202128074	31-MAY-2010 06:14	DONE	113	1.5	percent	100	75.0	83.3	117	125	8.33
989508	1202129297	04-JUN-2010 09:35	DONE	99	-0.12	percent	100	75.0	83.3	117	125	8.33
991377	1202133734	09-JUN-2010 01:01	DONE	114	1.6	percent	100	75.0	83.3	117	125	8.33
992256	1202135976	14-JUN-2010 06:19	DONE	105	0.66	percent	100	75.0	83.3	117	125	8.33
993157	1202138143	14-JUN-2010 06:26	DONE	114	1.6	percent	100	75.0	83.3	117	125	8.33
1000953	1202157540	11-JUL-2010 22:15	DONE	106	0.77	percent	100	75.0	83.3	117	125	8.33

Radon-222 RER: Limits LCL = 0 UCL = 3

Batch ID	Samp ID	Run Date	Status	Value	Deviation	Units	Mean	LCL	LWL	UWL	UCL	Stdev
942156	1202016820	15-JAN-2010 16:58	DONE	1.9	2	dec	0.53	0	-0.866	1.94	3.00	0.7
972884	1202089944	10-MAR-2010 01:04	DONE	0.23	-0.44	dec	0.53	0	-0.866	1.94	3.00	0.7
963652	1202067388	11-MAR-2010 11:58	DONE	0.31	-0.32	dec	0.53	0	-0.866	1.94	3.00	0.7
964261	1202068674	11-MAR-2010 22:44	DONE	0.09	-0.63	dec	0.53	0	-0.866	1.94	3.00	0.7
965705	1202072239	11-MAR-2010 22:44	DONE	0.09	-0.63	dec	0.53	0	-0.866	1.94	3.00	0.7
965846	1202072642	16-MAR-2010 21:33	DONE	0.26	-0.4	dec	0.53	0	-0.866	1.94	3.00	0.7
967258	1202076015	19-MAR-2010 23:08	DONE	0.41	-0.17	dec	0.53	0	-0.866	1.94	3.00	0.7
968303	1202078709	23-MAR-2010 20:44	DONE	0.05	-0.69	dec	0.53	0	-0.866	1.94	3.00	0.7
968842	1202079949	24-MAR-2010 21:06	DONE	0.63	0.13	dec	0.53	0	-0.866	1.94	3.00	0.7
970475	1202083970	31-MAR-2010 00:09	DONE	00	-0.76	dec	0.53	0	-0.866	1.94	3.00	0.7
1000953	1202157538	09-JUL-2010 17:13	DONE	1.9	2	dec	0.53	0	-0.866	1.94	3.00	0.7

Radon-222 SPIKE: Limits LCL = 75 UCL = 125

Batch ID	Samp ID	Run Date	Status	Value	Deviation	Units	Mean	LCL	LWL	UWL	UCL	Stdev
967258	1202076022	22-MAR-2010 13:32	DONE	101	0.16	percent	100	75.0	83.3	117	125	8.33
966794	1202074886	24-MAR-2010 10:43	DONE	120	2.3	percent	100	75.0	83.3	117	125	8.33
965846	1202072644	25-MAR-2010 03:41	DONE	116	1.9	percent	100	75.0	83.3	117	125	8.33
968842	1202079951	27-MAR-2010 07:07	DONE	121	2.6	percent	100	75.0	83.3	117	125	8.33
969311	1202081135	29-MAR-2010 04:53	DONE	101	0.07	percent	100	75.0	83.3	117	125	8.33
968204	1202078480	29-MAR-2010 22:10	DONE	102	0.22	percent	100	75.0	83.3	117	125	8.33

969700	1202082148	29-MAR-2010 22:59	DONE	92	-1	percent	100	75.0	83.3	117	125	8.33
968961	1202080317	30-MAR-2010 00:21	DONE	96	-0.51	percent	100	75.0	83.3	117	125	8.33
968303	1202078715	30-MAR-2010 14:46	DONE	117	2	percent	100	75.0	83.3	117	125	8.33
970477	1202083975	02-APR-2010 03:53	DONE	104	0.49	percent	100	75.0	83.3	117	125	8.33
970475	1202083972	03-APR-2010 05:33	DONE	99	-0.14	percent	100	75.0	83.3	117	125	8.33
970839	1202084976	03-APR-2010 05:45	DONE	108	0.97	percent	100	75.0	83.3	117	125	8.33
970058	1202082957	03-APR-2010 09:33	DONE	109	1.1	percent	100	75.0	83.3	117	125	8.33
971267	1202086036	05-APR-2010 04:45	DONE	100	0.05	percent	100	75.0	83.3	117	125	8.33
972847	1202089824	11-APR-2010 14:07	DONE	103	0.3	percent	100	75.0	83.3	117	125	8.33
973312	1202090934	11-APR-2010 14:07	DONE	88	-1	percent	100	75.0	83.3	117	125	8.33
974603	1202094122	16-APR-2010 15:45	DONE	104	0.48	percent	100	75.0	83.3	117	125	8.33
975561	1202096483	19-APR-2010 09:54	DONE	105	0.65	percent	100	75.0	83.3	117	125	8.33
975286	1202095816	19-APR-2010 10:42	DONE	118	2.1	percent	100	75.0	83.3	117	125	8.33
976372	1202098556	23-APR-2010 07:45	DONE	98	-0.2	percent	100	75.0	83.3	117	125	8.33
976828	1202099679	24-APR-2010 07:08	DONE	100	0.04	percent	100	75.0	83.3	117	125	8.33
977678	1202101492	26-APR-2010 18:39	DONE	107	0.88	percent	100	75.0	83.3	117	125	8.33
977370	1202100758	27-APR-2010 10:15	DONE	102	0.28	percent	100	75.0	83.3	117	125	8.33
978565	1202103620	30-APR-2010 03:00	DONE	96	-0.43	percent	100	75.0	83.3	117	125	8.33
978981	1202104625	01-MAY-2010 16:55	DONE	103	0.3	percent	100	75.0	83.3	117	125	8.33
979717	1202106351	02-MAY-2010 18:19	DONE	97	-0.36	percent	100	75.0	83.3	117	125	8.33
980337	1202107712	06-MAY-2010 03:39	DONE	106	0.74	percent	100	75.0	83.3	117	125	8.33
981849	1202111219	10-MAY-2010 06:37	DONE	104	0.44	percent	100	75.0	83.3	117	125	8.33
983453	1202115459	17-MAY-2010 17:20	DONE	129	3.4	percent	100	75.0	83.3	117	125	8.33
984715	1202118129	18-MAY-2010 00:42	DONE	99	-0.16	percent	100	75.0	83.3	117	125	8.33
984139	1202117011	19-MAY-2010 05:25	DONE	109	1.1	percent	100	75.0	83.3	117	125	8.33
983787	1202116238	19-MAY-2010 05:57	DONE	101	0.06	percent	100	75.0	83.3	117	125	8.33
985996	1202121047	23-MAY-2010 21:25	DONE	110	1.2	percent	100	75.0	83.3	117	125	8.33
988956	1202128073	31-MAY-2010 05:58	DONE	112	1.4	percent	100	75.0	83.3	117	125	8.33
989508	1202129296	04-JUN-2010 09:19	DONE	101	0.07	percent	100	75.0	83.3	117	125	8.33
991377	1202133733	09-JUN-2010 00:45	DONE	112	1.5	percent	100	75.0	83.3	117	125	8.33
992256	1202135975	14-JUN-2010 06:02	DONE	102	0.2	percent	100	75.0	83.3	117	125	8.33
993157	1202138142	14-JUN-2010 06:10	DONE	112	1.4	percent	100	75.0	83.3	117	125	8.33
1000953	1202157539	11-JUL-2010 21:59	DONE	110	1.3	percent	100	75.0	83.3	117	125	8.33

Generated 07/14/2010

LCS Recovery Radon-222

SPIKE Recovery Radon-222

RER Radon-222

BLANK Result Radon-222

● Denotes Outlier

● Denotes Outlier

Standards Data

ANALYTICS

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 · U.S.A.

Phone (404) 352-8677
Fax (404) 352-2837

0638

CERTIFICATE OF CALIBRATION
Standard Radionuclide Source

67519-278

Ra-226 5 mL Liquid in Flame Sealed Vial

This standard radionuclide source was prepared gravimetrically from a calibrated master solution. The master solution was calibrated using a germanium gamma spectrometer system.

Radionuclide purity and calibration were checked using a germanium gamma spectrometer system. The nuclear decay rate and assay date for this source are given below.

Analytics maintains traceability to the National Institute of Standards and Technology through participation in a Measurements Assurance Program as described in USNRC Reg. Guide 4.15, Revision 1, February 1979.

ISOTOPE:	Ra-226
ACTIVITY (dps):	2.353 E4
HALF-LIFE:	1.600 E3 years
CALIBRATION DATE:	January 23, 2004 12:00 EST
RELATIVE EXPANDED UNCERTAINTY (k=2):	3.3%

Impurities: γ-impurities (other than decay products) <0.1%

5.01065 grams 0.1M HCl solution with 50 μg/g Ba carrier.

P O NUMBER 3231RD, Item 5

SOURCE PREPARED BY: M. D. Currie
M. D. Currie, Radiochemist

Q A APPROVED: 1/26/04

RC-S-037-037a

Information about your session

Standard Traceability Log Rad

Source Material Info	
Parent Code:	0638
Prepared By:	Amanda Fehr
Carrier Conc:	0.1M HCl
Reference Date:	01/23/2004
Ampoule Mass (g):	5.01065 g
Uncertainty:	+/- 3.3 %
LogBook No:	RC-S-037-037

A Solution Material Info			
Isotope:	Radium-226		
Prepared By:	Amanda Fehr		
Prep Date:	01/16/2006		
Verification Date:	03/18/2010		
Expiration Date:	03/18/2011		
Primary Code:	0638-A		
Dilution(mL):	100 mL		
Mass of Parent(g):	4.8398 g		
Density(g/mL):	1.0266	Balance ID:	38080204

Calculations Converting parent activity to dpm/mL|dpm/g

(Mass of parent(g)) * (Parm Activity (dps)) * (conversion dpm to dps) / (Ampoule Mass(g) *(Dilution Vol)) = Parent Activity (dpm/mL)
(Mass of parent(g)) * (Parm Activity (dps)) * (conversion dpm to dps) / Density / (Ampoule Mass (g) * (Dilution Vol)) = Parent Activity (dpm/g)
(4.8398 g) * (23530 dps) * (60 dpm/dps) / (5.01065 g * 100 mL) = 13636.6133 dpm/mL
(4.8398 g) * (23530 dps) * (60 dpm/dps) / (1.0266 g/mL)/ (5.01065 g * 100 mL) = 13282.9676 dpm/g

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date
01/17/2006	Amanda Fehr	2.1041	100	0638-B	279.0211 dpm/mL	01/17/2007	01/17/2008
07/17/2006	Mary Aders	2.1313	100	0638-C	282.6281 dpm/mL	07/26/2006	07/26/2007
03/28/2007	Daniel Roy	2.1025	100	0638-D	279.2744 dpm/ml	04/08/2007	04/08/2008
03/28/2007	Daniel Roy	45.468	250	0638-E	2415.7999 dpm/ml	04/08/2010	04/08/2011
12/18/2007	Daniel Roy	2.014	100	0638-F	267.519 dpm/ml	02/02/2009	02/02/2010
02/12/2008	Daniel Roy	.5004	100	0638-G	66.468 dpm/ml	03/18/2010	03/18/2011
07/23/2008	Daniel Roy	5.0607	250	0638-H	268.8845 dpm/ml	07/17/2009	07/17/2010
09/28/2009	Angela Johnson	4.052	200	0638-I	269.113 dpm/mL	09/28/2009	09/28/2010

Verification for Ra-226 Standard 0638-E 10

	Isotope	Value	Uncertainty
G. Ramsay			
4/8/2010	0638-E	1134.617	21.9642
	0638-E	1066.158	20.7372
	0638-E	1098.519	21.3216

Mean Value (Counting) =	1099.765	101.33	**Pass**
Stdev =	34.24640148		**Rule 3 (Pass/Fail)**

Target =	1085.28
Lower Limit =	1031.271994
Upper Limit =	1168.2576
Rule 1 Pass/Fail	**Pass**
Two sigma =	68.49280295
10 % of Mean =	109.9764797
Rule 2 (Pass/Fail)	**Pass**

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements
Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.
Rule 3 = The determined mean value shall be within 5% of the certificate value.

The analyst prepared three standard verification sources for Ra-226 source 0638-E by transferring portions of the degassed standard into tared glass liquid scintillation vials. Ten mL of DI Water and 10 mL mineral oil was added to each vial and the vials were shaken to mix. A Blank vial was prepared in a similar fashion using 10 mL of DI water and 10 mL of mineral oil. The standard verification vials and Background source were dark adapted for two hours and counted on LSC Green using source standard verification. Each verification source calculation was performed as follows:

Source dpm/g = (A - B)/(C)(D)
where:
A = Ver. source cpm,
B = BKG cpm,
C = System efficiency, (cpm/dpm), and
D = mass used for standard verification.

Reference RAD SOP M-001

4/12/10 4/9/10

Runlogs

Instrument Run Log

Instrument Type: LSC **Batch ID: 1000953**

Sample ID	Sample Type	Analyst	Instrument	Run Date	Status	Geometry	Calibration Date
255994001	SAMPLE	AXW2	LSCRED	09-JUL-10 12:00	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
255994002	SAMPLE	AXW2	LSCRED	09-JUL-10 12:31	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
256024001	SAMPLE	AXW2	LSCRED	09-JUL-10 13:03	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
256024002	SAMPLE	AXW2	LSCRED	09-JUL-10 13:34	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
256024003	SAMPLE	AXW2	LSCRED	09-JUL-10 14:05	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
256024004	SAMPLE	AXW2	LSCRED	09-JUL-10 14:37	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
256024005	SAMPLE	AXW2	LSCRED	09-JUL-10 15:08	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
256024006	SAMPLE	AXW2	LSCRED	09-JUL-10 15:39	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
256024007	SAMPLE	AXW2	LSCRED	09-JUL-10 16:11	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
1202157537	MB	AXW2	LSCRED	09-JUL-10 16:42	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
1202157538	DUP	AXW2	LSCRED	09-JUL-10 17:13	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
1202157539	MS	AXW2	LSCRED	11-JUL-10 21:59	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00
1202157540	LCS	AXW2	LSCRED	11-JUL-10 22:15	DONE	10ML MINERAL OIL/10ML SAMPLE	28-JUL-09 00:00